

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD100720\
 Data File : VD067163.D
 Acq On : 07 Oct 2020 20:08
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 08 08:31:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 15:19:17 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.454	0.403	11.2	79	0.00
3 P	Chloromethane	0.398	0.353	11.3	79	0.00
4 C	Vinyl Chloride	0.399	0.363	9.0#	78	0.00
5 T	Bromomethane	0.291	0.277	4.8	85	0.00
6 T	Chloroethane	0.243	0.227	6.6	79	0.00
7 T	Trichlorofluoromethane	0.887	0.786	11.4	77	0.00
8 T	Diethyl Ether	0.220	0.221	-0.5	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.501	0.452	9.8	78	0.00
10 T	Methyl Iodide	0.487	0.473	2.9	79	0.00
11 T	Tert butyl alcohol	0.041	0.027	34.1#	82	0.00
12 CM	1,1-Dichloroethene	0.471	0.432	8.3#	78	0.00
13 T	Acrolein	0.032	0.030	6.3	83	0.00
14 T	Allyl chloride	0.612	0.571	6.7	76	0.00
15 T	Acrylonitrile	0.081	0.080	1.2	82	0.00
16 T	Acetone	0.074	0.071	4.1	79	0.00
17 T	Carbon Disulfide	1.447	1.295	10.5	76	0.00
18 T	Methyl Acetate	0.185	0.195	-5.4	88	0.00
19 T	Methyl tert-butyl Ether	0.860	0.926	-7.7	85	0.00
20 T	Methylene Chloride	0.670	0.570	14.9	87	0.00
21 T	trans-1,2-Dichloroethene	0.536	0.519	3.2	81	0.00
22 T	Diisopropyl ether	1.118	1.117	0.1	79	0.00
23 T	Vinyl Acetate	0.633	0.627	0.9	75	0.00
24 P	1,1-Dichloroethane	0.863	0.851	1.4	83	0.00
25 T	2-Butanone	0.104	0.096	7.7	78	0.00
26 T	2,2-Dichloropropane	0.828	0.702	15.2	75	0.00
27 T	cis-1,2-Dichloroethene	0.573	0.589	-2.8	85	0.00
28 T	Bromochloromethane	0.333	0.330	0.9	78	0.00
29 T	Tetrahydrofuran	0.060	0.057	5.0	76	0.00
30 C	Chloroform	0.936	0.955	-2.0#	87	0.00
31 T	Cyclohexane	0.723	0.596	17.6	72	0.00
32 T	1,1,1-Trichloroethane	0.867	0.850	2.0	83	0.00
33 S	1,2-Dichloroethane-d4	0.482	0.464	3.7	78	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	87	0.00
35 S	Dibromofluoromethane	0.327	0.331	-1.2	82	0.00
36 T	1,1-Dichloropropene	0.466	0.425	8.8	78	0.00
37 T	Ethyl Acetate	0.161	0.149	7.5	78	0.00
38 T	Carbon Tetrachloride	0.504	0.491	2.6	82	0.00
39 T	Methylcyclohexane	0.493	0.454	7.9	74	0.00
40 TM	Benzene	1.294	1.270	1.9	83	0.00
41 T	Methacrylonitrile	0.087	0.086	1.1	78	0.01
42 TM	1,2-Dichloroethane	0.362	0.366	-1.1	86	0.00
43 T	Isopropyl Acetate	0.297	0.289	2.7	80	0.00
44 TM	Trichloroethene	0.387	0.382	1.3	85	0.00
45 C	1,2-Dichloropropane	0.304	0.307	-1.0#	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD100720\
 Data File : VD067163.D
 Acq On : 07 Oct 2020 20:08
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 08 08:31:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 01 15:19:17 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.175	0.183	-4.6	86	0.00
47 T	Bromodichloromethane	0.454	0.469	-3.3	87	0.00
48 T	Methyl methacrylate	0.152	0.141	7.2	72	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	79	0.00
50 S	Toluene-d8	1.193	1.154	3.3	77	0.00
51 T	4-Methyl-2-Pentanone	0.148	0.146	1.4	79	0.00
52 CM	Toluene	0.834	0.863	-3.5#	85	0.00
53 T	t-1,3-Dichloropropene	0.406	0.420	-3.4	84	0.00
54 T	cis-1,3-Dichloropropene	0.496	0.504	-1.6	85	0.00
55 T	1,1,2-Trichloroethane	0.243	0.250	-2.9	86	0.00
56 T	Ethyl methacrylate	0.251	0.279	-11.2	84	0.00
57 T	1,3-Dichloropropane	0.394	0.403	-2.3	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.140	0.143	-2.1	80	0.00
59 T	2-Hexanone	0.099	0.098	1.0	78	0.00
60 T	Dibromochloromethane	0.324	0.336	-3.7	85	0.00
61 T	1,2-Dibromoethane	0.233	0.237	-1.7	83	0.00
62 S	4-Bromofluorobenzene	0.405	0.416	-2.7	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.364	0.370	-1.6	90	0.00
65 PM	Chlorobenzene	0.994	1.000	-0.6	87	0.00
66 T	1,1,1,2-Tetrachloroethane	0.378	0.390	-3.2	89	0.00
67 C	Ethyl Benzene	1.704	1.740	-2.1#	85	0.00
68 T	m/p-Xylenes	0.664	0.672	-1.2	84	0.00
69 T	o-Xylene	0.594	0.620	-4.4	85	0.00
70 T	Styrene	1.020	1.071	-5.0	85	0.00
71 P	Bromoform	0.196	0.211	-7.7	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.395	3.355	1.2	83	0.00
74 T	N-amyl acetate	0.610	0.572	6.2	80	0.00
75 P	1,1,2,2-Tetrachloroethane	0.572	0.549	4.0	85	0.00
76 T	1,2,3-Trichloropropane	0.436	0.379	13.1	75	0.00
77 T	Bromobenzene	0.820	0.846	-3.2	92	0.00
78 T	n-propylbenzene	4.012	3.901	2.8	82	0.00
79 T	2-Chlorotoluene	2.296	2.268	1.2	85	0.00
80 T	1,3,5-Trimethylbenzene	2.868	2.923	-1.9	85	0.00
81 T	trans-1,4-Dichloro-2-butene	0.175	0.168	4.0	84	0.00
82 T	4-Chlorotoluene	2.442	2.415	1.1	87	0.00
83 T	tert-Butylbenzene	2.415	2.489	-3.1	88	0.00
84 T	1,2,4-Trimethylbenzene	2.863	2.960	-3.4	88	0.00
85 T	sec-Butylbenzene	3.402	3.313	2.6	82	0.00
86 T	p-Isopropyltoluene	3.154	3.114	1.3	84	0.00
87 T	1,3-Dichlorobenzene	1.597	1.566	1.9	88	0.00
88 T	1,4-Dichlorobenzene	1.597	1.581	1.0	89	0.00
89 T	n-Butylbenzene	2.882	2.762	4.2	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD100720\
Data File : VD067163.D
Acq On : 07 Oct 2020 20:08
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Oct 08 08:31:23 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100120S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 01 15:19:17 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.552	0.527	4.5	84	0.00
91 T	1,2-Dichlorobenzene	1.395	1.389	0.4	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.092	0.089	3.3	85	0.00
93 T	1,2,4-Trichlorobenzene	0.902	0.957	-6.1	90	0.00
94 T	Hexachlorobutadiene	0.554	0.540	2.5	86	0.00
95 T	Naphthalene	1.505	1.648	-9.5	88	0.00
96 T	1,2,3-Trichlorobenzene	0.783	0.835	-6.6	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6