

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101221\
 Data File : VD070676.D
 Acq On : 12 Oct 2021 13:41
 Operator : VA/SY
 Sample : M4163-01
 Misc : 5.13G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-04-100821

Quant Time: Oct 13 06:35:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	336918	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	651419	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	610134	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	259670	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	197296	49.76	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.52%
35) Dibromofluoromethane	7.909	113	214007	49.41	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.82%
50) Toluene-d8	10.338	98	828980	50.07	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.14%
62) 4-Bromofluorobenzene	12.626	95	255935	46.25	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	92.50%
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.255	105	31308	1.95	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101221\
Data File : VD070676.D
Acq On : 12 Oct 2021 13:41
Operator : VA/SY
Sample : M4163-01
Misc : 5.13G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-04-100821

Quant Time: Oct 13 06:35:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

