

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101221\
 Data File : VD070682.D
 Acq On : 12 Oct 2021 16:32
 Operator : VA/SY
 Sample : M4127-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CHRT-27080

Quant Time: Oct 13 06:37:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	340474	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	668822	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	622671	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	229351	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	213312	53.24	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.48%
35) Dibromofluoromethane	7.914	113	223159	50.19	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.38%
50) Toluene-d8	10.338	98	861469	50.68	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.36%
62) 4-Bromofluorobenzene	12.626	95	245594	43.23	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.46%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101221\
Data File : VD070682.D
Acq On : 12 Oct 2021 16:32
Operator : VA/SY
Sample : M4127-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CHRT-27080

Quant Time: Oct 13 06:37:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration

