

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101221\
 Data File : VD070683.D
 Acq On : 12 Oct 2021 17:01
 Operator : VA/SY
 Sample : M4127-03
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CHRT-27051

Quant Time: Oct 13 06:37:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 06 09:10:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	356844	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	695749	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	663623	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	254887	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	205885	49.03	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.06%
35) Dibromofluoromethane	7.908	113	224563	48.55	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.10%
50) Toluene-d8	10.332	98	878707	49.69	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.38%
62) 4-Bromofluorobenzene	12.620	95	290711	49.19	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.38%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101221\
Data File : VD070683.D
Acq On : 12 Oct 2021 17:01
Operator : VA/SY
Sample : M4127-03
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CHRT-27051

Quant Time: Oct 13 06:37:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D100521S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 06 09:10:31 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.973 min Scan# 1029
Delta R.T. -0.006 min
Lab File: VD070683.D
Acq: 12 Oct 2021 17:01




Instrument :
MSVOA_D
ClientSampleId :
CHRT-27051


Tgt Ion:168 Resp: 356844
Ion Ratio Lower Upper
168 100
99 59.9 51.4 77.0



#33
1,2-Dichloroethane-d4
Concen: 49.03 ug/l
RT: 8.326 min Scan# 1089
Delta R.T. -0.006 min
Lab File: VD070683.D
Acq: 12 Oct 2021 17:01

Tgt Ion: 65 Resp: 205885
Ion Ratio Lower Upper
65 100
67 54.8 0.0 110.2









#34

1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.861 min Scan# 1180
Delta R.T. -0.000 min
Lab File: VD070683.D
Acq: 12 Oct 2021 17:01

Instrument :
MSVOA_D
ClientSampleId :
CHRT-27051




Tgt Ion:	114	Resp:	695749
Ion	Ratio	Lower	Upper
114	100		
63	19.6	0.0	41.8
88	16.3	0.0	32.6




#35

Dibromofluoromethane
Concen: 48.55 ug/l
RT: 7.908 min Scan# 1018
Delta R.T. -0.006 min
Lab File: VD070683.D
Acq: 12 Oct 2021 17:01

Tgt Ion:	113	Resp:	224563
Ion	Ratio	Lower	Upper
113	100		
111	103.2	83.1	124.7
192	17.4	12.7	19.1









#50
Toluene-d8
Concen: 49.69 ug/l
RT: 10.332 min Scan# 1430
Delta R.T. -0.006 min
Lab File: VD070683.D
Acq: 12 Oct 2021 17:01




Instrument :
MSVOA_D
ClientSampleId :
CHRT-27051


Tgt Ion: 98 Resp: 878707
Ion Ratio Lower Upper
98 100
100 65.2 52.2 78.4



#62
4-Bromofluorobenzene
Concen: 49.19 ug/l
RT: 12.620 min Scan# 1819
Delta R.T. -0.006 min
Lab File: VD070683.D
Acq: 12 Oct 2021 17:01


Tgt Ion: 95 Resp: 290711
Ion Ratio Lower Upper
95 100
174 70.7 0.0 138.0
176 69.2 0.0 135.6








#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.638 min Scan# 1652
Delta R.T. -0.000 min
Lab File: VD070683.D
Acq: 12 Oct 2021 17:01

Instrument :
MSVOA_D
ClientSampleId :
CHRT-27051



Tgt Ion:117 Resp: 663623
Ion Ratio Lower Upper
117 100
82 56.4 45.4 68.2
119 31.1 24.7 37.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.561 min Scan# 1979
Delta R.T. -0.000 min
Lab File: VD070683.D
Acq: 12 Oct 2021 17:01

Tgt Ion:152 Resp: 254887
Ion Ratio Lower Upper
152 100
115 63.6 31.9 95.7
150 159.5 0.0 350.0

