

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101322\
 Data File : VD074523.D
 Acq On : 13 Oct 2022 12:21
 Operator : VA/SY
 Sample : VD1013SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1013SBL01

Quant Time: Oct 13 16:55:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 12 13:14:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	97310	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	168145	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	140816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	64965	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	46887	55.978	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	111.960%
35) Dibromofluoromethane	7.805	113	55649	55.232	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.460%
50) Toluene-d8	10.269	98	201314	52.220	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.440%
62) 4-Bromofluorobenzene	12.569	95	58479	49.395	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101322\
Data File : VD074523.D
Acq On : 13 Oct 2022 12:21
Operator : VA/SY
Sample : VD1013SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1013SBL01

Quant Time: Oct 13 16:55:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 12 13:14:00 2022
Response via : Initial Calibration

