

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101322\
 Data File : VD074535.D
 Acq On : 13 Oct 2022 18:23
 Operator : VA/SY
 Sample : N5142-10
 Misc : 6.92G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-9

Quant Time: Oct 14 01:32:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 12 13:14:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	8598	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	16435	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	17856	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	7319	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	8351	112.840	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	225.680%#
35) Dibromofluoromethane	7.805	113	7448	76.704	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	153.400%#
50) Toluene-d8	10.269	98	22638	59.818	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	119.640%
62) 4-Bromofluorobenzene	12.575	95	6835	60.271	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	120.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101322\
Data File : VD074535.D
Acq On : 13 Oct 2022 18:23
Operator : VA/SY
Sample : N5142-10
Misc : 6.92G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-9

Quant Time: Oct 14 01:32:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 12 13:14:00 2022
Response via : Initial Calibration

