

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101322\
 Data File : VD074536.D
 Acq On : 13 Oct 2022 18:51
 Operator : VA/SY
 Sample : N5142-13
 Misc : 6.72G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-8

Quant Time: Oct 14 01:32:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 12 13:14:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	2348	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.769	114	4332	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	3731	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	1302	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	1906	94.307	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 188.620%#	
35) Dibromofluoromethane	7.799	113	1509	58.285	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 116.580%	
50) Toluene-d8	10.269	98	5234	52.682	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.360%	
62) 4-Bromofluorobenzene	12.575	95	1405	45.649	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 91.300%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101322\
Data File : VD074536.D
Acq On : 13 Oct 2022 18:51
Operator : VA/SY
Sample : N5142-13
Misc : 6.72G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-8

Quant Time: Oct 14 01:32:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 12 13:14:00 2022
Response via : Initial Calibration

