

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD101420\
 Data File : VD067269.D
 Acq On : 14 Oct 2020 14:22
 Operator : VA/SY
 Sample : L4404-01
 Misc : 7.49G/5.00ml/MSVOA D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SHELL-L15-L14-NORTH

Quant Time: Oct 15 06:13:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D101320S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 14 03:30:30 2020
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	317664	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	525158	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	485054	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	205074	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	170718	53.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.28%	
35) Dibromofluoromethane	7.92	113	175685	50.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.54%	
50) Toluene-d8	10.34	98	621290	49.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.24%	
62) 4-Bromofluorobenzene	12.64	95	208828	47.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.96%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.99	85	385	Below Cal	#	44
5) Bromomethane	2.77	94	120	Below Cal	#	3
16) Acetone	4.16	43	7199	16.152	ug/l	94
20) Methylene Chloride	4.94	84	648	Below Cal	#	65
25) 2-Butanone	7.22	43	1323	2.604	ug/l	# 45
29) Tetrahydrofuran	7.58	42	1677	5.938	ug/l	# 26
31) Cyclohexane	7.98	56	6310	1.721	ug/l	# 14
36) 1,1-Dichloropropene	7.98	75	22988	5.165	ug/l	# 49
38) Carbon Tetrachloride	7.98	117	32920	5.787	ug/l	# 16
51) 4-Methyl-2-Pentanone	10.34	43	2522	1.946	ug/l	# 1
67) Ethyl Benzene	11.75	91	180753	11.336	ug/l	99
68) m/p-Xylenes	11.86	106	41333	6.732	ug/l	98
69) o-Xylene	12.18	106	32312	5.683	ug/l	98
73) Isopropylbenzene	12.48	105	53920	4.143	ug/l	99
76) 1,2,3-Trichloropropane	12.64	75	107690	75.766	ug/l	# 100
78) n-propylbenzene	12.83	91	20329	1.348	ug/l	96
80) 1,3,5-Trimethylbenzene	12.97	105	82006	7.251	ug/l	99
81) trans-1,4-Dichloro-2-buten	12.64	75	107690	171.814	ug/l	# 10
83) tert-Butylbenzene	13.28	119	29907	3.130	ug/l	# 67
84) 1,2,4-Trimethylbenzene	13.27	105	215745	19.300	ug/l	100
85) sec-Butylbenzene	13.27	105	215745	16.457	ug/l	# 55
86) p-Isopropyltoluene	13.52	119	12724	1.028	ug/l	91
89) n-Butylbenzene	13.78	91	104021	9.556	ug/l	# 44
90) Hexachloroethane	14.12	117	4055	1.878	ug/l	# 9
95) Naphthalene	15.41	128	226376	37.481	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD101420\
Data File : VD067269.D
Acq On : 14 Oct 2020 14:22
Operator : VA/SY
Sample : L4404-01
Misc : 7.49G/5.00ml/MSVOA D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SHELL-L15-L14-NORTH

Quant Time: Oct 15 06:13:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D101320S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 14 03:30:30 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.98 min Scan# 1030

Delta R.T. -0.01 min

Lab File: VD067269.D

Acq: 14 Oct 2020 14:22

Instrument :

MSVOA_D

ClientSampleId :


SHELL-L15-L14-NORTH

Tot Ion:168 Resp: 317664

Ion Ratio Lower Upper

168 100

99 56.1 45.9 68.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC

7.98

150000



100000

50000

0

7.90 8.00 8.10

Time-->



#2

Dichlorodifluoromethane

Concen: Below Cal

RT: 1.99 min Scan# 11

Delta R.T. 0.00 min

Lab File: VD067269.D


Acq: 14 Oct 2020 14:22

Tgt Ion: 85 Resp: 385

Ion Ratio Lower Upper

85 100

87 0.0 15.2 45.6#



Abundance Ion 84.90 (84.60 to 85.60): VDC

Ion 86.90 (86.60 to 87.60): VDC

1.99

500

400

300


200


100

0

1.96 1.98 2.00

Time-->





#5

Bromomethane

Concen: Below Cal

RT: 2.77 min Scan# 144

Delta R.T. -0.01 min

Lab File: VD067269.D

Acq: 14 Oct 2020 14:22

Instrument :

MSVOA_D

ClientSampleId :


SHELL-L15-L14-NORTH

Tot Ion: 94 Resp: 120

Ion Ratio Lower Upper


94 100

96 0.0 73.8 110.6#





Abundance Ion 93.90 (93.60 to 94.60): VDC

Ion 95.90 (95.60 to 96.60): VDC



Time-->



#16

Acetone

Concen: 16.152 ug/l

RT: 4.16 min Scan# 380

Delta R.T. 0.00 min

Lab File: VD067269.D


Acq: 14 Oct 2020 14:22

Tgt Ion: 43 Resp: 7199

Ion Ratio Lower Upper


43 100

58 35.8 26.1 39.1





Abundance Ion 43.00 (42.70 to 43.70): VDC

Ion 58.00 (57.70 to 58.70): VDC



Time-->









#20
Methylene Chloride
Concen: Below Cal
RT: 4.94 min Scan# 514
Delta R.T. -0.02 min
Lab File: VD067269.D
Acq: 14 Oct 2020 14:22




Instrument :
MSVOA_D
ClientSampleId :
SHELL-L15-L14-NORTH


Tot Ion: 84 Resp: 648
Ion Ratio Lower Upper
84 100
49 44.5 72.2 108.2#
51 0.0 23.8 35.8#
86 63.8 48.2 72.4



#25
2-Butanone
Concen: 2.604 ug/l
RT: 7.22 min Scan# 901
Delta R.T. 0.01 min
Lab File: VD067269.D
Acq: 14 Oct 2020 14:22

Tgt Ion: 43 Resp: 1323
Ion Ratio Lower Upper
43 100
72 0.0 23.7 35.5#





#29

Tetrahydrofuran

Concen: 5.938 ug/l

RT: 7.58 min Scan# 962

Delta R.T. 0.01 min

Lab File: VD067269.D

Acq: 14 Oct 2020 14:22

Instrument :

MSVOA_D

ClientSampleId :

SHELL-L15-L14-NORTH


Tot Ion: 42 Resp: 1677

Ion Ratio Lower Upper




42 100

72 0.0 43.2 64.8#

71 0.0 39.4 59.0#



Abundance Ion 42.00 (41.70 to 42.70): VDC
Ion 72.00 (71.70 to 72.70): VDC
Ion 71.00 (70.70 to 71.70): VDC



#31

Cyclohexane

Concen: 1.721 ug/l

RT: 7.98 min Scan# 1031

Delta R.T. 0.00 min

Lab File: VD067269.D

Acq: 14 Oct 2020 14:22


Tgt Ion: 56 Resp: 6310

Ion Ratio Lower Upper



56 100


69 151.1 29.8 44.6#

84 139.9 76.6 114.8#



Abundance Ion 56.00 (55.70 to 56.70): VDC
Ion 69.00 (68.70 to 69.70): VDC
Ion 84.00 (83.70 to 84.70): VDC









#33
 1,2-Dichloroethane-d4
 Concen: 53.642 ug/l
 RT: 8.34 min Scan# 1091
 Delta R.T. 0.01 min
 Lab File: VD067269.D
 Acq: 14 Oct 2020 14:22




Instrument :
 MSVOA_D
 ClientSampleId :
 SHELL-L15-L14-NORTH


Tot Ion: 65 Resp: 170718
 Ion Ratio Lower Upper
 65 100
 67 49.5 0.0 103.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.87 min Scan# 1181
 Delta R.T. 0.00 min
 Lab File: VD067269.D
 Acq: 14 Oct 2020 14:22

Tgt Ion: 114 Resp: 525158
 Ion Ratio Lower Upper
 114 100
 63 18.7 0.0 35.0
 88 15.1 0.0 30.6









#35
 Dibromofluoromethane
 Concen: 50.268 ug/l
 RT: 7.92 min Scan# 1020
 Delta R.T. 0.01 min
 Lab File: VD067269.D
 Acq: 14 Oct 2020 14:22




Instrument :
 MSVOA_D
 ClientSampleId :
 SHELL-L15-L14-NORTH


Tot Ion	Ratio	Lower	Upper
113	100		
111	101.1	81.6	122.4
192	20.5	17.0	25.4



#36
 1,1-Dichloropropene
 Concen: 5.165 ug/l
 RT: 7.98 min Scan# 1031
 Delta R.T. -0.13 min
 Lab File: VD067269.D
 Acq: 14 Oct 2020 14:22

Tgt Ion	Ratio	Lower	Upper
75	100		
110	10.7	19.6	58.7#
77	0.0	24.6	37.0#





#38

Carbon Tetrachloride

Concen: 5.787 ug/l

RT: 7.98 min Scan# 1031

Delta R.T. -0.11 min

Lab File: VD067269.D

Acq: 14 Oct 2020 14:22

Instrument :

MSVOA_D

ClientSampleId :

SHELL-L15-L14-NORTH


Tot Ion:117 Resp: 32920

Ion Ratio Lower Upper

117 100

119 4.1 75.0 112.4#




121 0.0 24.9 37.3#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#50

Toluene-d8

Concen: 49.122 ug/l

RT: 10.34 min Scan# 1432

Delta R.T. 0.00 min

Lab File: VD067269.D


Acq: 14 Oct 2020 14:22

Tgt Ion: 98 Resp: 621290

Ion Ratio Lower Upper



98 100


100 65.7 53.1 79.7



Abundance Ion 98.00 (97.70 to 98.70): V

Ion 100.00 (99.70 to 100.70): V









#51
4-Methyl-2-Pentanone
Concen: 1.946 ug/l
RT: 10.34 min Scan# 1432
Delta R.T. 0.11 min
Lab File: VD067269.D
Acq: 14 Oct 2020 14:22




Instrument :
MSVOA_D
ClientSampleId :
SHELL-L15-L14-NORTH


Tot Ion: 43 Resp: 2522
Ion Ratio Lower Upper
43 100
58 183.5 35.7 53.5#



#62
4-Bromofluorobenzene
Concen: 47.483 ug/l
RT: 12.64 min Scan# 1822
Delta R.T. 0.00 min
Lab File: VD067269.D
Acq: 14 Oct 2020 14:22

Tgt Ion: 95 Resp: 208828
Ion Ratio Lower Upper
95 100
174 84.8 0.0 175.2
176 81.5 0.0 168.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.65 min Scan# 1654
Delta R.T. 0.00 min
Lab File: VD067269.D
Acq: 14 Oct 2020 14:22

Instrument :
MSVOA_D
ClientSampleId :
SHELL-L15-L14-NORTH

Tot Ion:117 Resp: 485054
Ion Ratio Lower Upper
117 100
82 53.0 39.9 59.9
119 32.2 24.9 37.3




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VDC
Ion 118.90 (118.60 to 119.60): V






#67
Ethyl Benzene
Concen: 11.336 ug/l
RT: 11.75 min Scan# 1671
Delta R.T. 0.00 min
Lab File: VD067269.D
Acq: 14 Oct 2020 14:22

Tgt Ion: 91 Resp: 180753
Ion Ratio Lower Upper
91 100
106 31.0 25.4 38.0



Abundance Ion 91.00 (90.70 to 91.70): VDC
Ion 106.00 (105.70 to 106.70): V









#68
m/p-Xylenes
Concen: 6.732 ug/l
RT: 11.86 min Scan# 1689
Delta R.T. 0.00 min
Lab File: VD067269.D
Acq: 14 Oct 2020 14:22




Instrument :
MSVOA_D
ClientSampleId :
SHELL-L15-L14-NORTH


Tot Ion:106 Resp: 41333
Ion Ratio Lower Upper
106 100
91 202.3 163.8 245.8



#69
o-Xylene
Concen: 5.683 ug/l
RT: 12.18 min Scan# 1745
Delta R.T. 0.00 min
Lab File: VD067269.D
Acq: 14 Oct 2020 14:22

Tgt Ion:106 Resp: 32312
Ion Ratio Lower Upper
106 100
91 221.8 109.1 327.3






#72




1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.58 min Scan# 1983
Delta R.T. 0.00 min
Lab File: VD067269.D
Acq: 14 Oct 2020 14:22

Instrument :
MSVOA_D
ClientSampleId :
SHELL-L15-L14-NORTH

Tot Ion:152 Resp: 205074
Ion Ratio Lower Upper
152 100
115 58.4 29.3 87.9
150 158.5 0.0 339.0




Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#73


Isopropylbenzene
Concen: 4.143 ug/l
RT: 12.48 min Scan# 1796
Delta R.T. 0.00 min
Lab File: VD067269.D
Acq: 14 Oct 2020 14:22

Tgt Ion:105 Resp: 53920
Ion Ratio Lower Upper
105 100
120 25.7 13.3 39.8



Abundance Ion 105.00 (104.70 to 105.70): V
Ion 120.00 (119.70 to 120.70): V





#76

1,2,3-Trichloropropane

Concen: 75.766 ug/l

RT: 12.64 min Scan# 1822

Delta R.T. -0.15 min

Lab File: VD067269.D

Acq: 14 Oct 2020 14:22

Instrument :

MSVOA_D

ClientSampleId :


SHELL-L15-L14-NORTH

Tot Ion: 75 Resp: 107690

Ion Ratio Lower Upper




75 100

77 2.0 0.0 0.0#



Abundance Ion 74.90 (74.60 to 75.60): VDC

Ion 77.00 (76.70 to 77.70): VDC



#78

n-propylbenzene

Concen: 1.348 ug/l

RT: 12.83 min Scan# 1854

Delta R.T. 0.00 min

Lab File: VD067269.D


Acq: 14 Oct 2020 14:22

Tgt Ion: 91 Resp: 20329

Ion Ratio Lower Upper



91 100


120 25.9 11.9 35.9



Abundance Ion 91.00 (90.70 to 91.70): VDC

Ion 120.00 (119.70 to 120.70): VDC





#80

1,3,5-Trimethylbenzene

Concen: 7.251 ug/l

RT: 12.97 min Scan# 1878

Delta R.T. 0.00 min

Lab File: VD067269.D

Acq: 14 Oct 2020 14:22

Instrument :

MSVOA_D

ClientSampleId :





SHELL-L15-L14-NORTH

Tot Ion:105 Resp: 82006

Ion Ratio Lower Upper

105 100

120 48.9 24.3 72.8



#81

trans-1,4-Dichloro-2-butene

Concen: 171.814 ug/l

RT: 12.64 min Scan# 1822

Delta R.T. 0.11 min

Lab File: VD067269.D

Acq: 14 Oct 2020 14:22




Tgt Ion: 75 Resp: 107690


Ion Ratio Lower Upper

75 100

53 0.1 78.5 117.7#

89 0.1 39.9 59.9#









#83
tert-Butylbenzene
Concen: 3.130 ug/l
RT: 13.28 min Scan# 1931
Delta R.T. 0.05 min
Lab File: VD067269.D
Acq: 14 Oct 2020 14:22




Instrument :
MSVOA_D
ClientSampleId :
SHELL-L15-L14-NORTH


Tot Ion:119 Resp: 29907
Ion Ratio Lower Upper
119 100
91 86.4 33.1 99.2
134 0.8 13.3 39.9#



#84
1,2,4-Trimethylbenzene
Concen: 19.300 ug/l
RT: 13.27 min Scan# 1930
Delta R.T. 0.00 min
Lab File: VD067269.D
Acq: 14 Oct 2020 14:22

Tgt Ion:105 Resp: 215745
Ion Ratio Lower Upper
105 100
120 45.2 22.4 67.3





#85

sec-Butylbenzene

Concen: 16.457 ug/l

RT: 13.27 min Scan# 1930

Delta R.T. -0.14 min

Lab File: VD067269.D

Acq: 14 Oct 2020 14:22

Instrument :

MSVOA_D

ClientSampleId :


SHELL-L15-L14-NORTH

Tot Ion:105 Resp: 215745

Ion Ratio Lower Upper

105 100

134 0.1 10.5 31.5#



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V

13.27



150000

100000

50000

0

Time-->



#86

p-Isopropyltoluene

Concen: 1.028 ug/l

RT: 13.52 min Scan# 1972

Delta R.T. -0.01 min

Lab File: VD067269.D

Acq: 14 Oct 2020 14:22


Tgt Ion:119 Resp: 12724

Ion Ratio Lower Upper

119 100

134 32.3 13.1 39.3

91 26.8 11.9 35.9



Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): V

12000

10000

8000


6000


4000

2000

0

Time-->









#89
n-Butylbenzene
Concen: 9.556 ug/l
RT: 13.78 min Scan# 2017
Delta R.T. -0.06 min
Lab File: VD067269.D
Acq: 14 Oct 2020 14:22




Instrument :
MSVOA_D
ClientSampleId :
SHELL-L15-L14-NORTH


Tot Ion: 91 Resp: 104021
Ion Ratio Lower Upper
91 100
92 10.6 26.7 80.0#
134 2.3 14.2 42.6#



#90
Hexachloroethane
Concen: 1.878 ug/l
RT: 14.12 min Scan# 2074
Delta R.T. 0.00 min
Lab File: VD067269.D
Acq: 14 Oct 2020 14:22

Tgt Ion: 117 Resp: 4055
Ion Ratio Lower Upper
117 100
201 0.0 40.8 122.2#





#95
 Naphthalene
 Concen: 37.481 ug/l
 RT: 15.41 min Scan# 2293
 Delta R.T. -0.01 min
 Lab File: VD067269.D
 Acq: 14 Oct 2020 14:22

Instrument :
 MSVOA_D
 ClientSampleId :
 SHELL-L15-L14-NORTH

Tot Ion	Ratio	Lower	Upper
128	100		
127	12.9	10.4	15.6
129	11.2	8.6	12.8

