


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD101520\
Data File : VD067273.D
Acq On : 15 Oct 2020 10:09
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D101320S.M
Title : SW846 8260
Last Update : Wed Oct 14 03:30:30 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	34813	PASS
75	95	30	60	49.8	103624	PASS
95	95	100	100	100.0	208277	PASS
96	95	5	9	6.7	13971	PASS
173	174	0.00	2	0.9	1658	PASS
174	95	50	100	86.3	179669	PASS
175	174	5	9	8.1	14482	PASS
176	174	95	101	99.3	178368	PASS
177	176	5	9	6.9	12309	PASS