

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101822\
 Data File : VD074569.D
 Acq On : 18 Oct 2022 17:02
 Operator : VA/SY
 Sample : N5184-01
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EO-02-101822

Quant Time: Oct 19 02:47:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 12 13:14:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	6909	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	12826	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	14180	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	6467	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	7715	129.730	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	259.460%#
35) Dibromofluoromethane	7.805	113	6273	83.012	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	166.020%#
50) Toluene-d8	10.269	98	16131	54.768	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.540%
62) 4-Bromofluorobenzene	12.569	95	5534	62.760	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	125.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101822\
Data File : VD074569.D
Acq On : 18 Oct 2022 17:02
Operator : VA/SY
Sample : N5184-01
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EO-02-101822

Quant Time: Oct 19 02:47:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101122S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 12 13:14:00 2022
Response via : Initial Calibration

