

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101922\
 Data File : VD074587.D
 Acq On : 19 Oct 2022 19:02
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 20 06:49:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 20 06:41:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.388	0.385	0.8	100	0.00
3 P	Chloromethane	0.525	0.536	-2.1	100	0.00
4 C	Vinyl Chloride	0.581	0.602	-3.6#	99	0.00
5 T	Bromomethane	0.440	0.397	9.8	98	0.00
6 T	Chloroethane	0.399	0.417	-4.5	100	0.00
7 T	Trichlorofluoromethane	0.824	0.827	-0.4	100	0.00
8 T	Diethyl Ether	0.238	0.255	-7.1	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.531	0.545	-2.6	99	0.00
10 T	Methyl Iodide	0.605	0.699	-15.5	100	0.00
11 T	Tert butyl alcohol	0.064	0.042	34.4#	107	0.00
12 CM	1,1-Dichloroethene	0.531	0.550	-3.6#	99	0.00
13 T	Acrolein	0.041	0.039	4.9	92	0.00
14 T	Allyl chloride	0.574	0.567	1.2	91	0.01
15 T	Acrylonitrile	0.097	0.106	-9.3	99	0.00
16 T	Acetone	0.083	0.070	15.7	80	0.00
17 T	Carbon Disulfide	1.454	1.388	4.5	96	0.00
18 T	Methyl Acetate	0.252	0.242	4.0	99	0.01
19 T	Methyl tert-butyl Ether	1.038	1.150	-10.8	101	0.00
20 T	Methylene Chloride	0.882	0.701	20.5	97	0.00
21 T	trans-1,2-Dichloroethene	0.592	0.600	-1.4	99	0.00
22 T	Diisopropyl ether	1.158	1.259	-8.7	101	0.00
23 T	Vinyl Acetate	0.506	0.508	-0.4	95	0.00
24 P	1,1-Dichloroethane	0.898	0.937	-4.3	98	0.00
25 T	2-Butanone	0.111	0.109	1.8	95	0.00
26 T	2,2-Dichloropropane	0.859	0.828	3.6	97	0.00
27 T	cis-1,2-Dichloroethene	0.657	0.685	-4.3	98	0.00
28 T	Bromochloromethane	0.219	0.281	-28.3#	104	0.00
29 T	Tetrahydrofuran	0.061	0.067	-9.8	101	0.00
30 C	Chloroform	0.967	1.014	-4.9#	100	0.00
31 T	Cyclohexane	0.767	0.727	5.2	97	0.00
32 T	1,1,1-Trichloroethane	0.887	0.913	-2.9	100	0.00
33 S	1,2-Dichloroethane-d4	0.382	0.367	3.9	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.306	0.263	14.1	98	0.00
36 T	1,1-Dichloropropene	0.449	0.442	1.6	99	0.00
37 T	Ethyl Acetate	0.142	0.140	1.4	95	0.00
38 T	Carbon Tetrachloride	0.440	0.440	0.0	98	0.00
39 T	Methylcyclohexane	0.537	0.502	6.5	90	0.00
40 TM	Benzene	1.333	1.359	-2.0	98	0.00
41 T	Methacrylonitrile	0.076	0.078	-2.6	114	0.00
42 TM	1,2-Dichloroethane	0.308	0.317	-2.9	99	0.00
43 T	Isopropyl Acetate	0.263	0.278	-5.7	99	0.00
44 TM	Trichloroethene	0.391	0.382	2.3	96	0.00
45 C	1,2-Dichloropropane	0.293	0.283	3.4#	93	0.00
46 T	Dibromomethane	0.169	0.165	2.4	93	0.00
47 T	Bromodichloromethane	0.418	0.410	1.9	92	0.00
48 T	Methyl methacrylate	0.120	0.123	-2.5	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101922\
 Data File : VD074587.D
 Acq On : 19 Oct 2022 19:02
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 20 06:49:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 20 06:41:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	94	0.00
50 S	Toluene-d8	1.001	0.909	9.2	89	0.00
51 T	4-Methyl-2-Pentanone	0.131	0.134	-2.3	94	0.00
52 CM	Toluene	0.822	0.826	-0.5#	93	0.00
53 T	t-1,3-Dichloropropene	0.367	0.355	3.3	90	0.00
54 T	cis-1,3-Dichloropropene	0.463	0.439	5.2	90	0.00
55 T	1,1,2-Trichloroethane	0.232	0.237	-2.2	95	0.00
56 T	Ethyl methacrylate	0.254	0.263	-3.5	93	0.00
57 T	1,3-Dichloropropane	0.376	0.374	0.5	93	0.00
58 T	2-Chloroethyl Vinyl ether	0.056	0.050	10.7	87	0.00
59 T	2-Hexanone	0.089	0.088	1.1	91	0.00
60 T	Dibromochloromethane	0.280	0.278	0.7	91	0.00
61 T	1,2-Dibromoethane	0.212	0.214	-0.9	93	0.00
62 S	4-Bromofluorobenzene	0.331	0.267	19.3	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.370	0.363	1.9	92	0.00
65 PM	Chlorobenzene	1.053	1.112	-5.6	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.362	0.389	-7.5	95	0.00
67 C	Ethyl Benzene	1.834	1.929	-5.2#	96	0.00
68 T	m/p-Xylenes	0.729	0.761	-4.4	102	0.00
69 T	o-Xylene	0.679	0.715	-5.3	100	0.00
70 T	Styrene	1.143	1.242	-8.7	103	0.00
71 P	Bromoform	0.190	0.202	-6.3	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.899	4.014	-2.9	101	0.00
74 T	N-amyl acetate	0.576	0.620	-7.6	104	0.00
75 P	1,1,2,2-Tetrachloroethane	0.617	0.642	-4.1	103	0.00
76 T	1,2,3-Trichloropropane	0.437	0.498	-14.0	102	0.00
77 T	Bromobenzene	0.880	0.913	-3.8	100	0.00
78 T	n-propylbenzene	4.608	4.769	-3.5	102	0.00
79 T	2-Chlorotoluene	2.497	2.537	-1.6	101	0.00
80 T	1,3,5-Trimethylbenzene	3.165	3.232	-2.1	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.171	0.168	1.8	103	0.00
82 T	4-Chlorotoluene	2.565	2.621	-2.2	103	0.00
83 T	tert-Butylbenzene	2.824	2.953	-4.6	104	0.00
84 T	1,2,4-Trimethylbenzene	3.115	3.209	-3.0	100	0.00
85 T	sec-Butylbenzene	4.238	4.358	-2.8	101	0.00
86 T	p-Isopropyltoluene	3.550	3.627	-2.2	100	0.00
87 T	1,3-Dichlorobenzene	1.829	1.850	-1.1	102	0.00
88 T	1,4-Dichlorobenzene	1.790	1.819	-1.6	103	0.00
89 T	n-Butylbenzene	3.234	3.340	-3.3	101	0.00
90 T	Hexachloroethane	0.618	0.626	-1.3	102	0.00
91 T	1,2-Dichlorobenzene	1.526	1.606	-5.2	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.085	0.092	-8.2	107	0.00
93 T	1,2,4-Trichlorobenzene	1.006	1.074	-6.8	106	0.00
94 T	Hexachlorobutadiene	0.554	0.573	-3.4	103	0.00
95 T	Naphthalene	1.894	2.011	-6.2	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101922\
Data File : VD074587.D
Acq On : 19 Oct 2022 19:02
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Oct 20 06:49:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 20 06:41:20 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.863	0.931	-7.9	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6