



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD101922\
Data File : VD074571.D
Acq On : 19 Oct 2022 09:20
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
Title : SW846 8260
Last Update : Thu Oct 20 06:41:20 2022



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	8065	PASS
75	95	30	60	49.8	24296	PASS
95	95	100	100	100.0	48749	PASS
96	95	5	9	6.5	3167	PASS
173	174	0.00	2	1.0	487	PASS
174	95	50	100	97.1	47317	PASS
175	174	5	9	8.2	3860	PASS
176	174	95	101	95.7	45269	PASS
177	176	5	9	6.4	2918	PASS