

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102021\
 Data File : VD070785.D
 Acq On : 20 Oct 2021 19:39
 Operator : VA/SY
 Sample : M4285-13
 Misc : 6.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MH-GB-(1)

Quant Time: Oct 21 05:41:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	46486	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	76372	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	87263	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	37460	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	51361	93.28	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	186.56%#
35) Dibromofluoromethane	7.914	113	34129	69.79	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	139.58%
50) Toluene-d8	10.332	98	101999	54.41	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.82%
62) 4-Bromofluorobenzene	12.620	95	35704	55.32	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.64%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102021\
Data File : VD070785.D
Acq On : 20 Oct 2021 19:39
Operator : VA/SY
Sample : M4285-13
Misc : 6.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MH-GB-(1)

Quant Time: Oct 21 05:41:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

