

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102022\
 Data File : VD074593.D
 Acq On : 20 Oct 2022 13:11
 Operator : VA/SY
 Sample : N5142-07
 Misc : 6.55G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-10

Quant Time: Oct 20 14:56:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 20 06:41:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	67804	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	115091	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	90879	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	39509	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	12056	23.289	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	46.580%#
35) Dibromofluoromethane	7.810	113	5591	7.840	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	15.680%#
50) Toluene-d8	10.269	98	4429	1.923	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	3.840%#
62) 4-Bromofluorobenzene	12.569	95	3759	3.886	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	7.780%#
Target Compounds						
16) Acetone	4.022	43	3180	28.131	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102022\
Data File : VD074593.D
Acq On : 20 Oct 2022 13:11
Operator : VA/SY
Sample : N5142-07
Misc : 6.55G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-10

Quant Time: Oct 20 14:56:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 20 06:41:20 2022
Response via : Initial Calibration

