

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102022\
 Data File : VD074594.D
 Acq On : 20 Oct 2022 13:39
 Operator : VA/SY
 Sample : N5142-13
 Misc : 6.65G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-8

Quant Time: Oct 20 14:56:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 20 06:41:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	52811	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	92895	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	80677	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	36944	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	11068	27.450	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	54.900%
35) Dibromofluoromethane	7.804	113	6574	11.935	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	23.860%#
50) Toluene-d8	10.269	98	11314	6.085	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	12.180%#
62) 4-Bromofluorobenzene	12.575	95	3684	5.150	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	10.300%#
Target Compounds						
16) Acetone	4.022	43	13221	150.159	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102022\
Data File : VD074594.D
Acq On : 20 Oct 2022 13:39
Operator : VA/SY
Sample : N5142-13
Misc : 6.65G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-8

Quant Time: Oct 20 14:56:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 20 06:41:20 2022
Response via : Initial Calibration

