

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102022\
 Data File : VD074595.D
 Acq On : 20 Oct 2022 14:06
 Operator : VA/SY
 Sample : N5142-04
 Misc : 6.37G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-11

Quant Time: Oct 20 14:57:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 20 06:41:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	104038	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	180173	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	147023	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	61582	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	49139	61.864	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.720%
35) Dibromofluoromethane	7.804	113	60818	61.172	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	122.340%
50) Toluene-d8	10.269	98	223620	62.010	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	124.020%
62) 4-Bromofluorobenzene	12.575	95	57060	55.202	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102022\
Data File : VD074595.D
Acq On : 20 Oct 2022 14:06
Operator : VA/SY
Sample : N5142-04
Misc : 6.37G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-11

Quant Time: Oct 20 14:57:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 20 06:41:20 2022
Response via : Initial Calibration

