

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102022\
 Data File : VD074597.D
 Acq On : 20 Oct 2022 15:01
 Operator : VA/SY
 Sample : N5184-01RE
 Misc : 5.10G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EO-02-101822RE

Quant Time: Oct 21 02:52:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 20 06:41:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	4375	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	8234	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	7929	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	3626	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	4246	127.117	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 254.240%#	
35) Dibromofluoromethane	7.805	113	3600	79.565	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 159.120%#	
50) Toluene-d8	10.263	98	9633	58.451	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 116.900%	
62) 4-Bromofluorobenzene	12.569	95	3220	68.638	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 137.280%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102022\
Data File : VD074597.D
Acq On : 20 Oct 2022 15:01
Operator : VA/SY
Sample : N5184-01RE
Misc : 5.10G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EO-02-101822RE

Quant Time: Oct 21 02:52:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 20 06:41:20 2022
Response via : Initial Calibration

