

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102022\
 Data File : VD074600.D
 Acq On : 20 Oct 2022 16:23
 Operator : VA/SY
 Sample : N5209-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CL-02-101922

Quant Time: Oct 21 03:08:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 20 06:41:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	86475	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	150227	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	127667	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	55837	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	46185	69.954	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	139.900%
35) Dibromofluoromethane	7.804	113	55062	66.519	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	133.040%
50) Toluene-d8	10.269	98	173450	57.685	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	115.380%
62) 4-Bromofluorobenzene	12.569	95	48566	56.392	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102022\
Data File : VD074600.D
Acq On : 20 Oct 2022 16:23
Operator : VA/SY
Sample : N5209-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CL-02-101922

Quant Time: Oct 21 03:08:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 20 06:41:20 2022
Response via : Initial Calibration

