

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102022\
 Data File : VD074601.D
 Acq On : 20 Oct 2022 16:51
 Operator : VA/SY
 Sample : N5219-02
 Misc : 6.35G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BACKYARD-BOTTOM-A

Quant Time: Oct 21 03:08:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 20 06:41:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	98673	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	166160	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	127563	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	48577	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	44173	58.635	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.280%
35) Dibromofluoromethane	7.805	113	56661	61.809	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	123.620%
50) Toluene-d8	10.269	98	189791	57.067	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	114.140%
62) 4-Bromofluorobenzene	12.569	95	46921	49.003	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102022\
Data File : VD074601.D
Acq On : 20 Oct 2022 16:51
Operator : VA/SY
Sample : N5219-02
Misc : 6.35G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BACKYARD-BOTTOM-A

Quant Time: Oct 21 03:08:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101922S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 20 06:41:20 2022
Response via : Initial Calibration

