

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD102120\
 Data File : VD067352.D
 Acq On : 21 Oct 2020 17:16
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 22 01:05:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D101320S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 19 19:13:29 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	92	0.00
2 T	Dichlorodifluoromethane	0.454	0.372	18.1	84	0.00
3 P	Chloromethane	0.315	0.283	10.2	91	0.00
4 C	Vinyl Chloride	0.345	0.300	13.0#	85	0.00
5 T	Bromomethane	0.267	0.238	10.9	91	0.00
6 T	Chloroethane	0.215	0.192	10.7	90	0.00
7 T	Trichlorofluoromethane	0.869	0.827	4.8	94	0.00
8 T	Diethyl Ether	0.199	0.202	-1.5	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.462	0.447	3.2	96	0.00
10 T	Methyl Iodide	0.426	0.401	5.9	85	0.00
11 T	Tert butyl alcohol	0.036	0.026	27.8#	100	-0.01
12 CM	1,1-Dichloroethene	0.413	0.402	2.7#	94	0.00
13 T	Acrolein	0.026	0.031	-19.2	96	0.00
14 T	Allyl chloride	0.473	0.454	4.0	92	0.00
15 T	Acrylonitrile	0.066	0.071	-7.6	106	0.01
16 T	Acetone	0.070	0.069	1.4	100	0.00
17 T	Carbon Disulfide	1.176	1.050	10.7	88	0.00
18 T	Methyl Acetate	0.160	0.166	-3.8	112	0.00
19 T	Methyl tert-butyl Ether	0.821	0.875	-6.6	102	0.00
20 T	Methylene Chloride	0.529	0.450	14.9	93	0.00
21 T	trans-1,2-Dichloroethene	0.475	0.469	1.3	97	0.00
22 T	Diisopropyl ether	0.863	0.891	-3.2	97	0.00
23 T	Vinyl Acetate	0.491	0.546	-11.2	101	0.00
24 P	1,1-Dichloroethane	0.739	0.727	1.6	96	0.00
25 T	2-Butanone	0.080	0.087	-8.7	108	0.00
26 T	2,2-Dichloropropane	0.762	0.727	4.6	94	0.00
27 T	cis-1,2-Dichloroethene	0.520	0.539	-3.7	100	0.00
28 T	Bromochloromethane	0.252	0.260	-3.2	92	0.00
29 T	Tetrahydrofuran	0.044	0.050	-13.6	108	0.00
30 C	Chloroform	0.883	0.899	-1.8#	98	0.00
31 T	Cyclohexane	0.577	0.508	12.0	90	0.00
32 T	1,1,1-Trichloroethane	0.866	0.856	1.2	96	0.00
33 S	1,2-Dichloroethane-d4	0.501	0.518	-3.4	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.333	0.347	-4.2	93	0.00
36 T	1,1-Dichloropropene	0.424	0.417	1.7	95	0.00
37 T	Ethyl Acetate	0.125	0.135	-8.0	104	0.00
38 T	Carbon Tetrachloride	0.542	0.542	0.0	95	0.00
39 T	Methylcyclohexane	0.443	0.433	2.3	91	0.00
40 TM	Benzene	1.163	1.178	-1.3	97	0.00
41 T	Methacrylonitrile	0.067	0.084	-25.4#	140	0.01
42 TM	1,2-Dichloroethane	0.364	0.374	-2.7	99	0.00
43 T	Isopropyl Acetate	0.247	0.268	-8.5	107	0.00
44 TM	Trichloroethene	0.372	0.369	0.8	95	0.00
45 C	1,2-Dichloropropane	0.269	0.273	-1.5#	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD102120\
 Data File : VD067352.D
 Acq On : 21 Oct 2020 17:16
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 22 01:05:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D101320S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 19 19:13:29 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.168	0.174	-3.6	104	0.00
47 T	Bromodichloromethane	0.450	0.459	-2.0	99	0.00
48 T	Methyl methacrylate	0.120	0.132	-10.0	109	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	111	0.00
50 S	Toluene-d8	1.204	1.228	-2.0	90	0.00
51 T	4-Methyl-2-Pentanone	0.123	0.136	-10.6	107	0.00
52 CM	Toluene	0.792	0.959	-21.1#	115	0.00
53 T	t-1,3-Dichloropropene	0.395	0.419	-6.1	101	0.00
54 T	cis-1,3-Dichloropropene	0.462	0.481	-4.1	100	0.00
55 T	1,1,2-Trichloroethane	0.229	0.238	-3.9	103	0.00
56 T	Ethyl methacrylate	0.237	0.270	-13.9	108	0.00
57 T	1,3-Dichloropropane	0.364	0.389	-6.9	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.122	0.132	-8.2	100	0.00
59 T	2-Hexanone	0.084	0.094	-11.9	105	0.00
60 T	Dibromochloromethane	0.334	0.349	-4.5	103	0.00
61 T	1,2-Dibromoethane	0.226	0.237	-4.9	104	0.00
62 S	4-Bromofluorobenzene	0.419	0.448	-6.9	97	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.362	0.349	3.6	97	0.00
65 PM	Chlorobenzene	0.959	0.962	-0.3	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.381	0.395	-3.7	100	0.00
67 C	Ethyl Benzene	1.644	1.687	-2.6#	98	0.00
68 T	m/p-Xylenes	0.633	0.648	-2.4	98	0.00
69 T	o-Xylene	0.586	0.606	-3.4	101	0.00
70 T	Styrene	1.009	1.045	-3.6	99	0.00
71 P	Bromoform	0.205	0.211	-2.9	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.173	3.312	-4.4	99	0.00
74 T	N-amyl acetate	0.479	0.514	-7.3	107	0.00
75 P	1,1,2,2-Tetrachloroethane	0.486	0.505	-3.9	105	0.00
76 T	1,2,3-Trichloropropane	0.347	0.379	-9.2	122	0.00
77 T	Bromobenzene	0.791	0.786	0.6	98	0.00
78 T	n-propylbenzene	3.677	3.748	-1.9	96	0.00
79 T	2-Chlorotoluene	2.132	2.122	0.5	97	0.00
80 T	1,3,5-Trimethylbenzene	2.758	2.837	-2.9	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.153	0.153	0.0	104	0.00
82 T	4-Chlorotoluene	2.263	2.291	-1.2	99	0.00
83 T	tert-Butylbenzene	2.329	2.412	-3.6	97	0.00
84 T	1,2,4-Trimethylbenzene	2.726	2.828	-3.7	98	0.00
85 T	sec-Butylbenzene	3.196	3.235	-1.2	97	0.00
86 T	p-Isopropyltoluene	3.017	3.049	-1.1	97	0.00
87 T	1,3-Dichlorobenzene	1.523	1.489	2.2	100	0.00
88 T	1,4-Dichlorobenzene	1.509	1.528	-1.3	103	0.00
89 T	n-Butylbenzene	2.654	2.719	-2.4	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD102120\
Data File : VD067352.D
Acq On : 21 Oct 2020 17:16
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Oct 22 01:05:43 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D101320S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 19 19:13:29 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.526	0.521	1.0	98	0.00
91 T	1,2-Dichlorobenzene	1.330	1.345	-1.1	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.088	0.088	0.0	108	0.00
93 T	1,2,4-Trichlorobenzene	0.897	0.968	-7.9	106	0.00
94 T	Hexachlorobutadiene	0.557	0.577	-3.6	100	0.00
95 T	Naphthalene	1.473	1.663	-12.9	108	0.00
96 T	1,2,3-Trichlorobenzene	0.775	0.843	-8.8	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6