

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102121\
 Data File : VD070803.D
 Acq On : 21 Oct 2021 16:15
 Operator : VA/SY
 Sample : M4297-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-04-10202021

Quant Time: Oct 22 06:15:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	16305	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	25268	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	26536	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	10558	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	17145	88.778	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 177.560%#	
35) Dibromofluoromethane	7.908	113	8918	55.120	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.240%	
50) Toluene-d8	10.332	98	32083	51.730	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.460%	
62) 4-Bromofluorobenzene	12.620	95	10983	51.429	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 102.860%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102121\
Data File : VD070803.D
Acq On : 21 Oct 2021 16:15
Operator : VA/SY
Sample : M4297-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-04-10202021

Quant Time: Oct 22 06:15:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

