

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102121\
 Data File : VD070808.D
 Acq On : 21 Oct 2021 18:44
 Operator : VA/SY
 Sample : M4294-10
 Misc : 5.53G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-2-GD

Quant Time: Oct 22 06:17:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	437972	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	806950	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	768425	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	321371	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	286968	55.319	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.640%
35) Dibromofluoromethane	7.909	113	265701	51.423	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.840%
50) Toluene-d8	10.332	98	1054945	53.262	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.520%
62) 4-Bromofluorobenzene	12.626	95	329785	48.355	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102121\
Data File : VD070808.D
Acq On : 21 Oct 2021 18:44
Operator : VA/SY
Sample : M4294-10
Misc : 5.53G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-2-GD

Quant Time: Oct 22 06:17:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

