

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102121\
 Data File : VD070811.D
 Acq On : 21 Oct 2021 20:10
 Operator : VA/SY
 Sample : M4290-01
 Misc : 5.32G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-1

Quant Time: Oct 22 06:18:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 18 08:32:03 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	428764	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	792829	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	750160	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	321343	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	269035	52.976	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.960%	
35) Dibromofluoromethane	7.909	113	260261	51.267	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.540%	
50) Toluene-d8	10.332	98	1029797	52.918	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.840%	
62) 4-Bromofluorobenzene	12.620	95	315312	47.057	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 94.120%	
Target Compounds						
16) Acetone	4.162	43	44220	36.102	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD102121\
Data File : VD070811.D
Acq On : 21 Oct 2021 20:10
Operator : VA/SY
Sample : M4290-01
Misc : 5.32G/5.00ml/MSVOA_D/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-1

Quant Time: Oct 22 06:18:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D101521S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 18 08:32:03 2021
Response via : Initial Calibration

