

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD102720\
 Data File : VD067414.D
 Acq On : 27 Oct 2020 16:24
 Operator : VA/SY
 Sample : L4528-03
 Misc : 5.80G/5.00ml/MSVOA D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CHRT-24245

Quant Time: Oct 28 01:25:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D101320S.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 19 19:13:29 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	104987	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	174283	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	142515	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	30866	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	77559	73.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	147.48%	
35) Dibromofluoromethane	7.92	113	65822	56.75	ug/l	0.00
Spiked Amount	50.000		Recovery	=	113.50%	
50) Toluene-d8	10.34	98	198050	47.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.36%	
62) 4-Bromofluorobenzene	12.64	95	46526	31.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	63.76%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.27	105	1707	1.015	ug/l	Qvalue 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD102720\
Data File : VD067414.D
Acq On : 27 Oct 2020 16:24
Operator : VA/SY
Sample : L4528-03
Misc : 5.80G/5.00ml/MSVOA D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CHRT-24245

Quant Time: Oct 28 01:25:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D101320S.M
Quant Title : SW846 8260
QLast Update : Mon Oct 19 19:13:29 2020
Response via : Initial Calibration

