


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD102720\
 Data File : VD067402.D
 Acq On : 27 Oct 2020 10:11
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D101320S.M
 Title : SW846 8260
 Last Update : Mon Oct 19 19:13:29 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	36578	PASS
75	95	30	60	50.4	112056	PASS
95	95	100	100	100.0	222357	PASS
96	95	5	9	6.2	13887	PASS
173	174	0.00	2	0.4	808	PASS
174	95	50	100	86.4	192021	PASS
175	174	5	9	7.8	15041	PASS
176	174	95	101	97.2	186666	PASS
177	176	5	9	6.8	12672	PASS