

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD102918\
 Data File : VD060238.D
 Acq On : 29 Oct 2018 19:30
 Operator : VA/AP
 Sample : J5194-09
 Misc : 5.94g/5ml/MSVOA D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HR-05-102618-B

Quant Time: Oct 30 01:44:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D100118S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 02 06:56:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.42	168	1480193	50.00	ug/l	0.03
34) 1,4-Difluorobenzene	7.45	114	2056883	50.00	ug/l	0.03
63) Chlorobenzene-d5	11.29	117	1403847	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	13.38	152	431167	50.00	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.79	65	652834	58.18	ug/l	0.03
Spiked Amount 50.000			Recovery	=	116.36%	
35) Dibromofluoromethane	6.33	113	854299	54.69	ug/l	0.03
Spiked Amount 50.000			Recovery	=	109.38%	
50) Toluene-d8	9.47	98	2132968	47.89	ug/l	0.03
Spiked Amount 50.000			Recovery	=	95.78%	
62) 4-Bromofluorobenzene	12.43	95	590382	35.40	ug/l	0.01
Spiked Amount 50.000			Recovery	=	70.80%	
Target Compounds						
8) Diethyl Ether	2.84	74	4334	1.713	ug/l	88
16) Acetone	3.17	43	15647	7.685	ug/l	99
20) Methylene Chloride	3.73	84	86396	1.193	ug/l	98
52) Toluene	9.56	92	151923	4.587	ug/l	100
68) m/p-Xylenes	11.57	106	30568	1.663	ug/l	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD102918\
Data File : VD060238.D
Acq On : 29 Oct 2018 19:30
Operator : VA/AP
Sample : J5194-09
Misc : 5.94g/5ml/MSVOA D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HR-05-102618-B

Quant Time: Oct 30 01:44:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D100118S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 02 06:56:35 2018
Response via : Initial Calibration

