


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD103020\
Data File : VD067460.D
Acq On : 30 Oct 2020 09:12
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D101320S.M
Title : SW846 8260
Last Update : Mon Oct 19 19:13:29 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.0	37032	PASS
75	95	30	60	51.3	118578	PASS
95	95	100	100	100.0	231210	PASS
96	95	5	9	6.5	14964	PASS
173	174	0.00	2	1.0	2041	PASS
174	95	50	100	87.7	202709	PASS
175	174	5	9	7.7	15537	PASS
176	174	95	101	96.4	195370	PASS
177	176	5	9	6.6	12814	PASS