

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD103119\
 Data File : VD064126.D
 Acq On : 31 Oct 2019 11:52
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 01 08:16:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D102919S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 30 08:00:10 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	0.511	0.491	3.9	102	0.00
3 P	Chloromethane	0.852	0.774	9.2	98	0.00
4 C	Vinyl Chloride	0.924	0.872	5.6#	98	0.00
5 T	Bromomethane	0.602	0.563	6.5	95	0.00
6 T	Chloroethane	0.574	0.527	8.2	96	0.00
7 T	Trichlorofluoromethane	1.272	1.213	4.6	99	0.00
8 T	Diethyl Ether	0.221	0.203	8.1	101	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.458	0.448	2.2	103	0.00
10 T	Methyl Iodide	0.568	0.539	5.1	92	0.00
11 T	Tert butyl alcohol	0.021	0.019	9.5	90	0.01
12 CM	1,1-Dichloroethene	0.434	0.413	4.8#	102	0.00
13 T	Acrolein	0.025	0.022	12.0	90	-0.01
14 T	Allyl chloride	0.508	0.491	3.3	100	0.00
15 T	Acrylonitrile	0.082	0.076	7.3	96	0.00
16 T	Acetone	0.076	0.083	-9.2	101	0.00
17 T	Carbon Disulfide	1.333	1.278	4.1	100	-0.01
18 T	Methyl Acetate	0.201	0.183	9.0	97	0.00
19 T	Methyl tert-butyl Ether	0.880	0.815	7.4	97	0.00
20 T	Methylene Chloride	0.541	0.476	12.0	102	0.00
21 T	trans-1,2-Dichloroethene	0.487	0.471	3.3	102	0.00
22 T	Diisopropyl ether	1.124	1.068	5.0	98	0.00
23 T	Vinyl Acetate	0.616	0.609	1.1	98	0.00
24 P	1,1-Dichloroethane	0.759	0.708	6.7	97	0.00
25 T	2-Butanone	0.105	0.098	6.7	94	0.00
26 T	2,2-Dichloropropane	0.721	0.680	5.7	99	0.00
27 T	cis-1,2-Dichloroethene	0.545	0.511	6.2	99	-0.01
28 T	Bromochloromethane	0.277	0.282	-1.8	98	-0.01
29 T	Tetrahydrofuran	0.061	0.057	6.6	94	0.00
30 C	Chloroform	0.869	0.807	7.1#	99	0.00
31 T	Cyclohexane	0.740	0.655	11.5	99	0.00
32 T	1,1,1-Trichloroethane	0.809	0.785	3.0	103	0.00
33 S	1,2-Dichloroethane-d4	0.398	0.394	1.0	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
35 S	Dibromofluoromethane	0.308	0.306	0.6	98	0.00
36 T	1,1-Dichloropropene	0.469	0.458	2.3	102	0.00
37 T	Ethyl Acetate	0.144	0.144	0.0	99	0.00
38 T	Carbon Tetrachloride	0.511	0.509	0.4	100	0.00
39 T	Methylcyclohexane	0.575	0.571	0.7	96	0.00
40 TM	Benzene	1.278	1.231	3.7	97	0.00
41 T	Methacrylonitrile	0.083	0.085	-2.4	103	0.00
42 TM	1,2-Dichloroethane	0.337	0.326	3.3	95	0.00
43 T	Isopropyl Acetate	0.291	0.286	1.7	99	0.00
44 TM	Trichloroethene	0.387	0.390	-0.8	100	0.00
45 C	1,2-Dichloropropane	0.296	0.287	3.0#	99	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD103119\
 Data File : VD064126.D
 Acq On : 31 Oct 2019 11:52
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 01 08:16:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D102919S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 30 08:00:10 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.178	0.173	2.8	97	0.00
47 T	Bromodichloromethane	0.450	0.450	0.0	98	0.00
48 T	Methyl methacrylate	0.135	0.142	-5.2	104	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	93	0.00
50 S	Toluene-d8	1.230	1.243	-1.1	102	0.00
51 T	4-Methyl-2-Pentanone	0.159	0.157	1.3	96	0.00
52 CM	Toluene	0.895	0.894	0.1#	99	0.00
53 T	t-1,3-Dichloropropene	0.420	0.423	-0.7	96	0.00
54 T	cis-1,3-Dichloropropene	0.494	0.490	0.8	97	0.00
55 T	1,1,2-Trichloroethane	0.253	0.235	7.1	91	0.00
56 T	Ethyl methacrylate	0.275	0.281	-2.2	97	0.00
57 T	1,3-Dichloropropane	0.418	0.410	1.9	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.111	0.120	-8.1	110	0.00
59 T	2-Hexanone	0.113	0.114	-0.9	93	0.00
60 T	Dibromochloromethane	0.343	0.346	-0.9	99	0.00
61 T	1,2-Dibromoethane	0.253	0.251	0.8	96	0.00
62 S	4-Bromofluorobenzene	0.408	0.413	-1.2	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.359	0.348	3.1	97	0.00
65 PM	Chlorobenzene	1.033	1.001	3.1	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.387	0.365	5.7	95	0.00
67 C	Ethyl Benzene	1.838	1.805	1.8#	98	0.00
68 T	m/p-Xylenes	0.739	0.739	0.0	98	0.00
69 T	o-Xylene	0.675	0.683	-1.2	99	0.00
70 T	Styrene	1.163	1.171	-0.7	96	0.00
71 P	Bromoform	0.206	0.202	1.9	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.474	3.346	3.7	98	0.00
74 T	N-amyl acetate	0.577	0.551	4.5	94	0.00
75 P	1,1,2,2-Tetrachloroethane	0.548	0.486	11.3	93	0.00
76 T	1,2,3-Trichloropropane	0.336	0.356	-6.0	99	0.00
77 T	Bromobenzene	0.839	0.786	6.3	96	0.00
78 T	n-propylbenzene	4.072	3.986	2.1	99	0.00
79 T	2-Chlorotoluene	2.183	2.070	5.2	97	0.00
80 T	1,3,5-Trimethylbenzene	2.908	2.820	3.0	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.165	0.166	-0.6	100	0.00
82 T	4-Chlorotoluene	2.332	2.248	3.6	98	0.00
83 T	tert-Butylbenzene	2.612	2.544	2.6	97	0.00
84 T	1,2,4-Trimethylbenzene	2.921	2.911	0.3	100	0.00
85 T	sec-Butylbenzene	3.587	3.585	0.1	98	0.00
86 T	p-Isopropyltoluene	3.370	3.400	-0.9	98	0.00
87 T	1,3-Dichlorobenzene	1.681	1.644	2.2	98	0.00
88 T	1,4-Dichlorobenzene	1.657	1.617	2.4	99	0.00
89 T	n-Butylbenzene	3.108	3.123	-0.5	97	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD103119\
Data File : VD064126.D
Acq On : 31 Oct 2019 11:52
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Nov 01 08:16:48 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D102919S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 30 08:00:10 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.631	0.593	6.0	95	0.00
91 T	1,2-Dichlorobenzene	1.447	1.396	3.5	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.080	0.072	10.0	95	0.00
93 T	1,2,4-Trichlorobenzene	1.040	1.015	2.4	97	0.00
94 T	Hexachlorobutadiene	0.608	0.611	-0.5	99	0.00
95 T	Naphthalene	1.696	1.643	3.1	95	0.00
96 T	1,2,3-Trichlorobenzene	0.901	0.859	4.7	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6