

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD110520\
 Data File : VD067537.D
 Acq On : 05 Nov 2020 11:53
 Operator : VA/SY
 Sample : L4590-01
 Misc : 4.67G/5.00ml/MSVOA D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NB-08-110320

Quant Time: Nov 06 00:34:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110320S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 00:56:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	312888	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	514314	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	507876	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	265475	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	160495	58.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.32%	
35) Dibromofluoromethane	7.92	113	160006	51.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.28%	
50) Toluene-d8	10.34	98	592748	54.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.86%	
62) 4-Bromofluorobenzene	12.64	95	220866	58.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.50%	
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.27	105	52257	3.582	ug/l	95
95) Naphthalene	15.40	128	23437	3.216	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD110520\
Data File : VD067537.D
Acq On : 05 Nov 2020 11:53
Operator : VA/SY
Sample : L4590-01
Misc : 4.67G/5.00ml/MSVOA D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NB-08-110320

Quant Time: Nov 06 00:34:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110320S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 00:56:28 2020
Response via : Initial Calibration

