

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD110520\
 Data File : VD067539.D
 Acq On : 05 Nov 2020 16:13
 Operator : VA/SY
 Sample : L4653-01
 Misc : 6.15G/5.00ml/MSVOA D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-7-COMP

Quant Time: Nov 06 00:38:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110320S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 00:56:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	279425	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	453899	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	459094	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	235339	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.33	65	147186	60.24	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	120.48%	
35) Dibromofluoromethane	7.91	113	149571	54.69	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	109.38%	
50) Toluene-d8	10.34	98	530486	55.20	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	110.40%	
62) 4-Bromofluorobenzene	12.64	95	202169	60.93	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	121.86%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD110520\
Data File : VD067539.D
Acq On : 05 Nov 2020 16:13
Operator : VA/SY
Sample : L4653-01
Misc : 6.15G/5.00ml/MSVOA D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-7-COMP

Quant Time: Nov 06 00:38:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110320S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 00:56:28 2020
Response via : Initial Calibration

