


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD110920\
Data File : VD067554.D
Acq On : 09 Nov 2020 09:27
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110320S.M
Title : SW846 8260
Last Update : Wed Nov 04 00:56:28 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.8	35832	PASS
75	95	30	60	52.2	111522	PASS
95	95	100	100	100.0	213738	PASS
96	95	5	9	6.8	14492	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	88.5	189098	PASS
175	174	5	9	7.4	13938	PASS
176	174	95	101	98.4	186112	PASS
177	176	5	9	6.7	12420	PASS