

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD111620\
 Data File : VD067644.D
 Acq On : 16 Nov 2020 16:46
 Operator : VA/SY
 Sample : L4687-08
 Misc : 4.42G/5.01ml/MSVOA D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SS-7

Quant Time: Nov 17 01:40:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110320S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 00:56:28 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	253609	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	410421	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	428247	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	217331	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	122149	55.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.16%	
35) Dibromofluoromethane	7.92	113	128253	51.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.74%	
50) Toluene-d8	10.34	98	481485	55.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.82%	
62) 4-Bromofluorobenzene	12.64	95	183292	61.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	122.18%	
Target Compounds						
27) cis-1,2-Dichloroethene	7.21	96	53184	20.783	ug/l	89
44) Trichloroethene	9.12	130	130461	41.694	ug/l	93
64) Tetrachloroethene	10.88	164	462017	145.675	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD111620\
Data File : VD067644.D
Acq On : 16 Nov 2020 16:46
Operator : VA/SY
Sample : L4687-08
Misc : 4.42G/5.01ml/MSVOA D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SS-7

Quant Time: Nov 17 01:40:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D110320S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 00:56:28 2020
Response via : Initial Calibration

