

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD111620\
 Data File : VD067630.D
 Acq On : 16 Nov 2020 10:11
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 17 01:12:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110320S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 00:56:28 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.457	0.409	10.5	85	0.00
3 P	Chloromethane	0.311	0.255	18.0	79	0.00
4 C	Vinyl Chloride	0.347	0.302	13.0#	78	0.00
5 T	Bromomethane	0.281	0.254	9.6	82	0.00
6 T	Chloroethane	0.214	0.201	6.1	81	0.00
7 T	Trichlorofluoromethane	0.899	0.916	-1.9	92	0.00
8 T	Diethyl Ether	0.181	0.173	4.4	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.454	0.456	-0.4	89	0.00
10 T	Methyl Iodide	0.486	0.487	-0.2	80	0.00
11 T	Tert butyl alcohol	0.022	0.021	4.5	95	-0.02
12 CM	1,1-Dichloroethene	0.417	0.405	2.9#	87	0.00
13 T	Acrolein	0.016	0.019	-18.7	83	-0.01
14 T	Allyl chloride	0.447	0.440	1.6	85	0.00
15 T	Acrylonitrile	0.067	0.064	4.5	82	-0.01
16 T	Acetone	0.065	0.069	-6.2	82	0.00
17 T	Carbon Disulfide	1.279	1.180	7.7	82	0.00
18 T	Methyl Acetate	0.135	0.125	7.4	89	0.00
19 T	Methyl tert-butyl Ether	0.823	0.842	-2.3	87	0.00
20 T	Methylene Chloride	0.462	0.431	6.7	87	0.00
21 T	trans-1,2-Dichloroethene	0.486	0.478	1.6	87	0.00
22 T	Diisopropyl ether	0.870	0.866	0.5	85	0.00
23 T	Vinyl Acetate	0.456	0.470	-3.1	84	0.00
24 P	1,1-Dichloroethane	0.712	0.703	1.3	88	0.00
25 T	2-Butanone	0.070	0.066	5.7	81	0.00
26 T	2,2-Dichloropropane	0.751	0.793	-5.6	92	0.00
27 T	cis-1,2-Dichloroethene	0.505	0.519	-2.8	88	0.00
28 T	Bromochloromethane	0.253	0.240	5.1	80	0.00
29 T	Tetrahydrofuran	0.042	0.041	2.4	83	0.00
30 C	Chloroform	0.849	0.873	-2.8#	90	0.00
31 T	Cyclohexane	0.634	0.578	8.8	83	-0.01
32 T	1,1,1-Trichloroethane	0.876	0.900	-2.7	92	0.00
33 S	1,2-Dichloroethane-d4	0.437	0.412	5.7	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
35 S	Dibromofluoromethane	0.301	0.304	-1.0	87	0.00
36 T	1,1-Dichloropropene	0.428	0.454	-6.1	88	0.00
37 T	Ethyl Acetate	0.111	0.111	0.0	78	0.00
38 T	Carbon Tetrachloride	0.556	0.624	-12.2	93	0.00
39 T	Methylcyclohexane	0.504	0.548	-8.7	85	0.00
40 TM	Benzene	1.159	1.236	-6.6	88	0.00
41 T	Methacrylonitrile	0.066	0.059	10.6	80	-0.01
42 TM	1,2-Dichloroethane	0.358	0.390	-8.9	91	0.00
43 T	Isopropyl Acetate	0.220	0.225	-2.3	85	0.00
44 TM	Trichloroethene	0.381	0.401	-5.2	89	0.00
45 C	1,2-Dichloropropane	0.251	0.263	-4.8#	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD111620\
 Data File : VD067630.D
 Acq On : 16 Nov 2020 10:11
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 17 01:12:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110320S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 00:56:28 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.159	0.166	-4.4	86	0.00
47 T	Bromodichloromethane	0.433	0.474	-9.5	89	0.00
48 T	Methyl methacrylate	0.113	0.109	3.5	75	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	82	0.00
50 S	Toluene-d8	1.059	1.038	2.0	86	0.00
51 T	4-Methyl-2-Pentanone	0.105	0.113	-7.6	86	0.00
52 CM	Toluene	0.777	0.862	-10.9#	90	0.00
53 T	t-1,3-Dichloropropene	0.371	0.416	-12.1	91	0.00
54 T	cis-1,3-Dichloropropene	0.441	0.460	-4.3	86	0.00
55 T	1,1,2-Trichloroethane	0.211	0.224	-6.2	87	0.00
56 T	Ethyl methacrylate	0.218	0.236	-8.3	86	0.00
57 T	1,3-Dichloropropane	0.339	0.359	-5.9	86	0.00
58 T	2-Chloroethyl Vinyl ether	0.102	0.100	2.0	76	0.00
59 T	2-Hexanone	0.073	0.084	-15.1	87	0.00
60 T	Dibromochloromethane	0.312	0.336	-7.7	88	0.00
61 T	1,2-Dibromoethane	0.212	0.224	-5.7	88	0.00
62 S	4-Bromofluorobenzene	0.365	0.377	-3.3	90	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
64 T	Tetrachloroethene	0.370	0.391	-5.7	87	0.00
65 PM	Chlorobenzene	0.963	1.043	-8.3	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.410	-7.9	89	0.00
67 C	Ethyl Benzene	1.648	1.852	-12.4#	89	0.00
68 T	m/p-Xylenes	0.641	0.713	-11.2	89	0.00
69 T	o-Xylene	0.572	0.652	-14.0	90	0.00
70 T	Styrene	0.998	1.116	-11.8	89	0.00
71 P	Bromoform	0.194	0.213	-9.8	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
73 T	Isopropylbenzene	3.231	3.696	-14.4	90	0.00
74 T	N-amyl acetate	0.425	0.448	-5.4	84	0.00
75 P	1,1,2,2-Tetrachloroethane	0.441	0.462	-4.8	88	0.00
76 T	1,2,3-Trichloropropane	0.302	0.287	5.0	84	0.00
77 T	Bromobenzene	0.794	0.870	-9.6	90	0.00
78 T	n-propylbenzene	3.696	4.160	-12.6	89	0.00
79 T	2-Chlorotoluene	2.100	2.341	-11.5	89	0.00
80 T	1,3,5-Trimethylbenzene	2.760	3.192	-15.7	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.149	0.160	-7.4	93	0.00
82 T	4-Chlorotoluene	2.252	2.518	-11.8	90	0.00
83 T	tert-Butylbenzene	2.339	2.714	-16.0	91	0.00
84 T	1,2,4-Trimethylbenzene	2.748	3.111	-13.2	88	0.00
85 T	sec-Butylbenzene	3.158	3.601	-14.0	88	0.00
86 T	p-Isopropyltoluene	3.050	3.486	-14.3	90	0.00
87 T	1,3-Dichlorobenzene	1.514	1.689	-11.6	92	0.00
88 T	1,4-Dichlorobenzene	1.533	1.646	-7.4	90	0.00
89 T	n-Butylbenzene	2.634	3.018	-14.6	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD111620\
Data File : VD067630.D
Acq On : 16 Nov 2020 10:11
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Nov 17 01:12:17 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110320S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 00:56:28 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.574	0.642	-11.8	91	0.00
91 T	1,2-Dichlorobenzene	1.311	1.440	-9.8	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.078	0.088	-12.8	96	0.00
93 T	1,2,4-Trichlorobenzene	0.890	1.011	-13.6	91	0.00
94 T	Hexachlorobutadiene	0.595	0.655	-10.1	91	0.00
95 T	Naphthalene	1.373	1.567	-14.1	87	0.00
96 T	1,2,3-Trichlorobenzene	0.762	0.828	-8.7	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6