

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111825\
 Data File : VD080532.D
 Acq On : 18 Nov 2025 10:37
 Operator : RP/MD
 Sample : VD1118SBL01
 Misc : 5.00G/5ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1118SBL01

Quant Time: Nov 18 11:35:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111725S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 18 00:45:54 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.876	168	297421	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.776	114	489620	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.582	117	408840	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.517	152	177899	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.229	65	118300	47.371	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.740%
35) Dibromofluoromethane	7.806	113	140095	48.602	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.200%
50) Toluene-d8	10.270	98	559255	47.960	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.920%
62) 4-Bromofluorobenzene	12.570	95	159796	47.666	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD111825\
Data File : VD080532.D
Acq On : 18 Nov 2025 10:37
Operator : RP/MD
Sample : VD1118SBL01
Misc : 5.00G/5ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1118SBL01

Quant Time: Nov 18 11:35:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D111725S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 18 00:45:54 2025
Response via : Initial Calibration

