

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD111919\
 Data File : VD064280.D
 Acq On : 19 Nov 2019 10:19
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 20 03:57:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110819S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 12 02:43:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	70	0.00
2 T	Dichlorodifluoromethane	0.415	0.395	4.8	65	0.00
3 P	Chloromethane	0.445	0.445	0.0	69	0.00
4 C	Vinyl Chloride	0.476	0.540	-13.4#	76	0.00
5 T	Bromomethane	0.300	0.367	-22.3#	81	0.00
6 T	Chloroethane	0.304	0.365	-20.1#	80	0.00
7 T	Trichlorofluoromethane	0.771	0.947	-22.8#	82	0.00
8 T	Diethyl Ether	0.212	0.209	1.4	67	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.441	0.444	-0.7	69	-0.01
10 T	Methyl Iodide	0.440	0.472	-7.3	61	0.00
11 T	Tert butyl alcohol	0.029	0.028	3.4	63	0.00
12 CM	1,1-Dichloroethene	0.420	0.420	0.0	67	0.00
13 T	Acrolein	0.024	0.026	-8.3	63	0.00
14 T	Allyl chloride	0.743	0.677	8.9	60	0.00
15 T	Acrylonitrile	0.090	0.091	-1.1	68	0.00
16 T	Acetone	0.097	0.095	2.1	63	0.00
17 T	Carbon Disulfide	1.308	1.232	5.8	62	0.00
18 T	Methyl Acetate	0.272	0.262	3.7	68	0.00
19 T	Methyl tert-butyl Ether	1.003	0.968	3.5	64	0.00
20 T	Methylene Chloride	0.492	0.432	12.2	64	-0.01
21 T	trans-1,2-Dichloroethene	0.476	0.469	1.5	66	0.00
22 T	Diisopropyl ether	1.378	1.312	4.8	64	0.00
23 T	Vinyl Acetate	0.768	0.771	-0.4	65	0.00
24 P	1,1-Dichloroethane	0.800	0.784	2.0	65	0.00
25 T	2-Butanone	0.126	0.122	3.2	65	0.00
26 T	2,2-Dichloropropane	0.785	0.747	4.8	64	0.00
27 T	cis-1,2-Dichloroethene	0.510	0.515	-1.0	68	0.00
28 T	Bromochloromethane	0.279	0.300	-7.5	68	0.00
29 T	Tetrahydrofuran	0.076	0.074	2.6	65	0.00
30 C	Chloroform	0.819	0.823	-0.5#	67	0.00
31 T	Cyclohexane	0.807	0.742	8.1	65	0.00
32 T	1,1,1-Trichloroethane	0.784	0.775	1.1	67	0.00
33 S	1,2-Dichloroethane-d4	0.411	0.395	3.9	61	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	70	0.00
35 S	Dibromofluoromethane	0.287	0.282	1.7	62	0.00
36 T	1,1-Dichloropropene	0.461	0.461	0.0	66	0.00
37 T	Ethyl Acetate	0.195	0.189	3.1	66	0.00
38 T	Carbon Tetrachloride	0.494	0.503	-1.8	68	0.00
39 T	Methylcyclohexane	0.584	0.578	1.0	67	0.00
40 TM	Benzene	1.261	1.248	1.0	65	0.00
41 T	Methacrylonitrile	0.122	0.109	10.7	57	0.00
42 TM	1,2-Dichloroethane	0.379	0.363	4.2	64	0.00
43 T	Isopropyl Acetate	0.401	0.377	6.0	63	0.00
44 TM	Trichloroethene	0.379	0.379	0.0	67	0.00
45 C	1,2-Dichloropropane	0.307	0.303	1.3#	66	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD111919\
 Data File : VD064280.D
 Acq On : 19 Nov 2019 10:19
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 20 03:57:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110819S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 12 02:43:24 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.164	0.164	0.0	67	0.00
47 T	Bromodichloromethane	0.443	0.438	1.1	66	0.00
48 T	Methyl methacrylate	0.196	0.174	11.2	61	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	67	0.00
50 S	Toluene-d8	1.083	1.055	2.6	61	0.00
51 T	4-Methyl-2-Pentanone	0.192	0.184	4.2	65	0.00
52 CM	Toluene	0.829	0.810	2.3#	65	0.00
53 T	t-1,3-Dichloropropene	0.443	0.426	3.8	64	0.00
54 T	cis-1,3-Dichloropropene	0.517	0.500	3.3	64	0.00
55 T	1,1,2-Trichloroethane	0.223	0.225	-0.9	67	0.00
56 T	Ethyl methacrylate	0.312	0.303	2.9	65	0.00
57 T	1,3-Dichloropropane	0.389	0.390	-0.3	67	0.00
58 T	2-Chloroethyl Vinyl ether	0.126	0.125	0.8	67	0.00
59 T	2-Hexanone	0.137	0.129	5.8	63	0.00
60 T	Dibromochloromethane	0.311	0.314	-1.0	68	0.00
61 T	1,2-Dibromoethane	0.223	0.228	-2.2	68	0.00
62 S	4-Bromofluorobenzene	0.375	0.350	6.7	60	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	69	0.00
64 T	Tetrachloroethene	0.385	0.385	0.0	66	0.00
65 PM	Chlorobenzene	1.017	1.017	0.0	67	0.00
66 T	1,1,1,2-Tetrachloroethane	0.387	0.389	-0.5	67	0.00
67 C	Ethyl Benzene	1.839	1.845	-0.3#	67	0.00
68 T	m/p-Xylenes	0.705	0.705	0.0	66	0.00
69 T	o-Xylene	0.654	0.646	1.2	66	0.00
70 T	Styrene	1.141	1.131	0.9	66	0.00
71 P	Bromoform	0.220	0.222	-0.9	67	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	71	0.00
73 T	Isopropylbenzene	3.670	3.611	1.6	67	0.00
74 T	N-amyl acetate	0.867	0.805	7.2	64	0.00
75 P	1,1,2,2-Tetrachloroethane	0.553	0.545	1.4	68	0.00
76 T	1,2,3-Trichloropropane	0.403	0.380	5.7	64	0.01
77 T	Bromobenzene	0.892	0.875	1.9	66	0.00
78 T	n-propylbenzene	4.193	4.146	1.1	66	0.00
79 T	2-Chlorotoluene	2.367	2.288	3.3	65	0.00
80 T	1,3,5-Trimethylbenzene	3.016	2.915	3.3	66	0.00
81 T	trans-1,4-Dichloro-2-butene	0.203	0.193	4.9	65	0.00
82 T	4-Chlorotoluene	2.469	2.399	2.8	66	0.00
83 T	tert-Butylbenzene	2.659	2.564	3.6	65	0.00
84 T	1,2,4-Trimethylbenzene	2.977	2.922	1.8	66	0.00
85 T	sec-Butylbenzene	3.571	3.551	0.6	67	0.00
86 T	p-Isopropyltoluene	3.395	3.375	0.6	67	0.00
87 T	1,3-Dichlorobenzene	1.655	1.641	0.8	68	0.00
88 T	1,4-Dichlorobenzene	1.615	1.627	-0.7	67	0.00
89 T	n-Butylbenzene	3.068	3.082	-0.5	67	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD111919\
Data File : VD064280.D
Acq On : 19 Nov 2019 10:19
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Nov 20 03:57:12 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D110819S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 12 02:43:24 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.627	0.617	1.6	67	0.00
91 T	1,2-Dichlorobenzene	1.401	1.417	-1.1	68	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.099	0.089	10.1	63	0.00
93 T	1,2,4-Trichlorobenzene	1.081	1.089	-0.7	67	0.00
94 T	Hexachlorobutadiene	0.698	0.701	-0.4	68	0.00
95 T	Naphthalene	1.771	1.758	0.7	67	0.00
96 T	1,2,3-Trichlorobenzene	0.919	0.927	-0.9	68	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5