

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD113020\
 Data File : VD067819.D
 Acq On : 30 Nov 2020 17:27
 Operator : VA/SY
 Sample : L4880-07
 Misc : 4.86G/5.00ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 KENS-B4-112320-6-8

Quant Time: Dec 01 00:51:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 24 00:51:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.99	168	190494	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	282435	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	262365	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	130918	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	93593	50.53	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.06%
35) Dibromofluoromethane	7.92	113	85642	47.65	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.30%
50) Toluene-d8	10.34	98	294862	44.87	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.74%
62) 4-Bromofluorobenzene	12.64	95	98746	43.96	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.92%

Target Compounds					Qvalue
16) Acetone	4.16	43	6534	27.307 ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD113020\
Data File : VD067819.D
Acq On : 30 Nov 2020 17:27
Operator : VA/SY
Sample : L4880-07
Misc : 4.86G/5.00ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
KENS-B4-112320-6-8

Quant Time: Dec 01 00:51:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 24 00:51:35 2020
Response via : Initial Calibration

