

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD113020\
 Data File : VD067821.D
 Acq On : 30 Nov 2020 18:22
 Operator : VA/SY
 Sample : L4891-07
 Misc : 6.40G/5.00ml/MSVOA D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PAV-B3-112420-10.5-11.5

Quant Time: Dec 01 00:59:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 24 00:51:35 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	161799	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	246967	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	191802	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	62470	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	99321	63.13	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery	=	126.26%	
35) Dibromofluoromethane	7.92	113	82540	52.52	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery	=	105.04%	
50) Toluene-d8	10.34	98	251876	43.83	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery	=	87.66%	
62) 4-Bromofluorobenzene	12.63	95	60247	30.67	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery	=	61.34%	

Target Compounds					Qvalue
16) Acetone	4.16	43	3009	14.805 ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD113020\
Data File : VD067821.D
Acq On : 30 Nov 2020 18:22
Operator : VA/SY
Sample : L4891-07
Misc : 6.40G/5.00ml/MSVOA D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PAV-B3-112420-10.5-11.5

Quant Time: Dec 01 00:59:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D112320S.M
Quant Title : SW846 8260
QLast Update : Tue Nov 24 00:51:35 2020
Response via : Initial Calibration

