

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD120220\
 Data File : VD067850.D
 Acq On : 02 Dec 2020 21:52
 Operator : VA/SY
 Sample : L4911-03
 Misc : 6.40G/5.00ml/MSVOA D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SK-12-1-SPC

Quant Time: Dec 03 02:47:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 02 08:05:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	238540	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	341266	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	300018	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	143165	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	122340	47.77	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.54%
35) Dibromofluoromethane	7.92	113	104337	45.69	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.38%
50) Toluene-d8	10.34	98	385995	46.96	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	93.92%
62) 4-Bromofluorobenzene	12.64	95	120247	42.31	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.62%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD120220\
Data File : VD067850.D
Acq On : 02 Dec 2020 21:52
Operator : VA/SY
Sample : L4911-03
Misc : 6.40G/5.00ml/MSVOA D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SK-12-1-SPC

Quant Time: Dec 03 02:47:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 02 08:05:58 2020
Response via : Initial Calibration

