

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD120220\
 Data File : VD067835.D
 Acq On : 02 Dec 2020 14:39
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 03 01:34:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 02 08:05:58 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.435	0.463	-6.4	106	0.00
3 P	Chloromethane	0.251	0.208	17.1	87	0.00
4 C	Vinyl Chloride	0.316	0.314	0.6#	104	0.00
5 T	Bromomethane	0.282	0.418	-48.2#	161#	0.00
6 T	Chloroethane	0.206	0.210	-1.9	103	0.00
7 T	Trichlorofluoromethane	0.968	0.978	-1.0	108	0.00
8 T	Diethyl Ether	0.189	0.179	5.3	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.441	0.450	-2.0	107	0.00
10 T	Methyl Iodide	0.468	0.269	42.5#	56	0.00
11 T	Tert butyl alcohol	0.042	0.030	28.6#	113	0.02
12 CM	1,1-Dichloroethene	0.414	0.426	-2.9#	108	0.00
13 T	Acrolein	0.021	0.018	14.3	86	0.00
14 T	Allyl chloride	0.383	0.402	-5.0	106	0.00
15 T	Acrylonitrile	0.065	0.063	3.1	96	0.00
16 T	Acetone	0.066	0.075	-13.6	104	0.00
17 T	Carbon Disulfide	1.157	1.138	1.6	104	0.00
18 T	Methyl Acetate	0.135	0.120	11.1	97	0.00
19 T	Methyl tert-butyl Ether	0.887	0.914	-3.0	101	0.00
20 T	Methylene Chloride	0.504	0.444	11.9	103	0.00
21 T	trans-1,2-Dichloroethene	0.471	0.482	-2.3	106	0.00
22 T	Diisopropyl ether	0.771	0.772	-0.1	100	0.00
23 T	Vinyl Acetate	0.429	0.460	-7.2	100	0.00
24 P	1,1-Dichloroethane	0.676	0.668	1.2	100	0.00
25 T	2-Butanone	0.070	0.069	1.4	95	0.00
26 T	2,2-Dichloropropane	0.806	0.840	-4.2	110	0.00
27 T	cis-1,2-Dichloroethene	0.516	0.503	2.5	100	0.00
28 T	Bromochloromethane	0.215	0.210	2.3	104	0.00
29 T	Tetrahydrofuran	0.040	0.039	2.5	95	0.00
30 C	Chloroform	0.903	0.869	3.8#	100	0.00
31 T	Cyclohexane	0.566	0.555	1.9	109	0.00
32 T	1,1,1-Trichloroethane	0.953	0.966	-1.4	107	0.00
33 S	1,2-Dichloroethane-d4	0.537	0.513	4.5	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.335	0.326	2.7	97	0.00
36 T	1,1-Dichloropropene	0.433	0.435	-0.5	104	0.00
37 T	Ethyl Acetate	0.117	0.107	8.5	95	0.00
38 T	Carbon Tetrachloride	0.655	0.655	0.0	106	0.00
39 T	Methylcyclohexane	0.513	0.533	-3.9	107	0.00
40 TM	Benzene	1.184	1.155	2.4	102	0.00
41 T	Methacrylonitrile	0.059	0.054	8.5	85	0.00
42 TM	1,2-Dichloroethane	0.422	0.406	3.8	97	0.00
43 T	Isopropyl Acetate	0.230	0.223	3.0	97	0.00
44 TM	Trichloroethene	0.407	0.397	2.5	103	0.00
45 C	1,2-Dichloropropane	0.243	0.229	5.8#	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD120220\
 Data File : VD067835.D
 Acq On : 02 Dec 2020 14:39
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 03 01:34:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 02 08:05:58 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.169	0.160	5.3	97	0.00
47 T	Bromodichloromethane	0.482	0.461	4.4	101	0.00
48 T	Methyl methacrylate	0.117	0.124	-6.0	109	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	100	-0.01
50 S	Toluene-d8	1.204	1.205	-0.1	101	0.00
51 T	4-Methyl-2-Pentanone	0.109	0.105	3.7	96	0.00
52 CM	Toluene	0.828	0.809	2.3#	101	0.00
53 T	t-1,3-Dichloropropene	0.416	0.407	2.2	100	0.00
54 T	cis-1,3-Dichloropropene	0.452	0.433	4.2	98	0.00
55 T	1,1,2-Trichloroethane	0.231	0.211	8.7	98	0.00
56 T	Ethyl methacrylate	0.235	0.226	3.8	95	0.00
57 T	1,3-Dichloropropane	0.366	0.344	6.0	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.107	0.111	-3.7	96	0.00
59 T	2-Hexanone	0.077	0.077	0.0	96	0.00
60 T	Dibromochloromethane	0.349	0.330	5.4	100	0.00
61 T	1,2-Dibromoethane	0.227	0.212	6.6	99	0.00
62 S	4-Bromofluorobenzene	0.416	0.439	-5.5	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
64 T	Tetrachloroethene	0.408	0.385	5.6	105	0.00
65 PM	Chlorobenzene	1.032	1.000	3.1	104	0.00
66 T	1,1,1,2-Tetrachloroethane	0.421	0.401	4.8	103	0.00
67 C	Ethyl Benzene	1.757	1.755	0.1#	105	0.00
68 T	m/p-Xylenes	0.693	0.685	1.2	105	0.00
69 T	o-Xylene	0.615	0.614	0.2	103	0.00
70 T	Styrene	1.070	1.042	2.6	100	0.00
71 P	Bromoform	0.223	0.210	5.8	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
73 T	Isopropylbenzene	3.438	3.379	1.7	106	0.00
74 T	N-amyl acetate	0.425	0.408	4.0	101	0.00
75 P	1,1,2,2-Tetrachloroethane	0.454	0.401	11.7	97	0.00
76 T	1,2,3-Trichloropropane	0.364	0.324	11.0	102	0.00
77 T	Bromobenzene	0.855	0.798	6.7	100	0.00
78 T	n-propylbenzene	3.811	3.723	2.3	105	0.00
79 T	2-Chlorotoluene	2.242	2.125	5.2	103	0.00
80 T	1,3,5-Trimethylbenzene	2.977	2.914	2.1	105	0.00
81 T	trans-1,4-Dichloro-2-butene	0.152	0.130	14.5	93	0.00
82 T	4-Chlorotoluene	2.392	2.270	5.1	104	0.00
83 T	tert-Butylbenzene	2.533	2.545	-0.5	104	0.00
84 T	1,2,4-Trimethylbenzene	2.960	2.836	4.2	101	0.00
85 T	sec-Butylbenzene	3.381	3.285	2.8	106	0.00
86 T	p-Isopropyltoluene	3.314	3.152	4.9	102	0.00
87 T	1,3-Dichlorobenzene	1.649	1.504	8.8	103	0.00
88 T	1,4-Dichlorobenzene	1.650	1.491	9.6	103	0.00
89 T	n-Butylbenzene	2.753	2.710	1.6	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD120220\
Data File : VD067835.D
Acq On : 02 Dec 2020 14:39
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Dec 03 01:34:02 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 02 08:05:58 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.612	0.563	8.0	105	0.00
91 T	1,2-Dichlorobenzene	1.430	1.298	9.2	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.095	0.083	12.6	97	0.00
93 T	1,2,4-Trichlorobenzene	1.000	0.975	2.5	104	0.00
94 T	Hexachlorobutadiene	0.653	0.626	4.1	104	0.00
95 T	Naphthalene	1.628	1.579	3.0	99	0.00
96 T	1,2,3-Trichlorobenzene	0.864	0.816	5.6	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6