

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD120419\
 Data File : VD064415.D
 Acq On : 04 Dec 2019 14:24
 Operator : VA/SY
 Sample : K6119-02
 Misc : 4.21G/5.00ml/MSVOA D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-3-(8-10)

Quant Time: Dec 05 06:14:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 29 07:18:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	724551	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	1111194	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	1004997	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	493358	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.33	65	327730	49.97	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.94%	
35) Dibromofluoromethane	7.92	113	341689	48.81	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	97.62%	
50) Toluene-d8	10.34	98	1311988	49.75	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.50%	
62) 4-Bromofluorobenzene	12.64	95	434002	51.21	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	102.42%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD120419\
Data File : VD064415.D
Acq On : 04 Dec 2019 14:24
Operator : VA/SY
Sample : K6119-02
Misc : 4.21G/5.00ml/MSVOA D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-3-(8-10)

Quant Time: Dec 05 06:14:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D112719S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 29 07:18:03 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.98 min Scan# 1031

Delta R.T. 0.01 min

Lab File: VD064415.D

Acq: 04 Dec 2019 14:24

Instrument :

MSVOA_D

ClientSampleId :


SB-3-(8-10)

Tot Ion:168 Resp: 724551

Ion Ratio Lower Upper

168 100

99 49.6 40.7 61.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC

7.98



300000

200000

100000

0

Time-->



#33

1,2-Dichloroethane-d4

Concen: 49.972 ug/l

RT: 8.33 min Scan# 1090

Delta R.T. -0.01 min

Lab File: VD064415.D


Acq: 04 Dec 2019 14:24

Tgt Ion: 65 Resp: 327730

Ion Ratio Lower Upper

65 100

67 52.0 0.0 102.0



Abundance Ion 64.90 (64.60 to 65.60): VDC

Ion 66.90 (66.60 to 67.60): VDC

8.33


150000


100000

50000

0

Time-->









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.87 min Scan# 1181
 Delta R.T. 0.00 min
 Lab File: VD064415.D
 Acq: 04 Dec 2019 14:24

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-3-(8-10)

Tot Ion:114 Resp: 1111194
 Ion Ratio Lower Upper
 114 100
 63 19.6 0.0 40.0
 88 14.4 0.0 29.2




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): V
 Ion 87.90 (87.60 to 88.60): V






#35
 Dibromofluoromethane
 Concen: 48.805 ug/l
 RT: 7.92 min Scan# 1020
 Delta R.T. 0.00 min
 Lab File: VD064415.D
 Acq: 04 Dec 2019 14:24

Tgt Ion:113 Resp: 341689
 Ion Ratio Lower Upper
 113 100
 111 101.7 79.4 119.2
 192 22.9 18.6 28.0



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 49.749 ug/l

RT: 10.34 min Scan# 1432

Delta R.T. 0.00 min

Lab File: VD064415.D

Acq: 04 Dec 2019 14:24

Instrument :

MSVOA_D

ClientSampleId :


SB-3-(8-10)

Tot Ion: 98 Resp: 1311988

Ion Ratio Lower Upper

98 100

100 65.2 52.6 79.0



Abundance Ion 98.00 (97.70 to 98.70): VDC

Ion 100.00 (99.70 to 100.70): VDC

10.34

800000



600000

400000

200000

0

Time--> 10.20 10.30 10.40 10.50



#62

4-Bromofluorobenzene

Concen: 51.208 ug/l

RT: 12.64 min Scan# 1822

Delta R.T. 0.00 min

Lab File: VD064415.D

Acq: 04 Dec 2019 14:24


Tgt Ion: 95 Resp: 434002

Ion Ratio Lower Upper

95 100

174 95.9 0.0 197.2

176 92.6 0.0 189.6



Abundance Ion 94.90 (94.60 to 95.60): VDC

Ion 173.80 (173.50 to 174.50): VDC

Ion 175.80 (175.50 to 176.50): VDC

12.64


300000


200000

100000

0

Time--> 12.60 12.70









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.65 min Scan# 1654
Delta R.T. 0.00 min
Lab File: VD064415.D
Acq: 04 Dec 2019 14:24

Instrument :
MSVOA_D
ClientSampleId :
SB-3-(8-10)

Tot Ion:117 Resp: 1004997
Ion Ratio Lower Upper
117 100
82 52.9 41.4 62.2
119 32.8 26.2 39.2




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.58 min Scan# 1983
Delta R.T. 0.00 min
Lab File: VD064415.D
Acq: 04 Dec 2019 14:24

Tgt Ion:152 Resp: 493358
Ion Ratio Lower Upper
152 100
115 55.0 26.7 80.1
150 157.5 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

