

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD120720\
 Data File : VD067894.D
 Acq On : 07 Dec 2020 12:36
 Operator : VA/SY
 Sample : L4969-02
 Misc : 6.42G/5.00ml/MSVOA D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NORTH-BRANFORD-02

Quant Time: Dec 08 03:28:10 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 02 08:05:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	223925	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	306922	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	271469	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	131302	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.34	65	110607	46.01	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.02%
35) Dibromofluoromethane	7.92	113	91604	44.60	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.20%
50) Toluene-d8	10.34	98	343581	46.47	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.94%
62) 4-Bromofluorobenzene	12.64	95	107541	42.07	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.14%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD120720\
Data File : VD067894.D
Acq On : 07 Dec 2020 12:36
Operator : VA/SY
Sample : L4969-02
Misc : 6.42G/5.00ml/MSVOA D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NORTH-BRANFORD-02

Quant Time: Dec 08 03:28:10 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 02 08:05:58 2020
Response via : Initial Calibration

