

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD120720\
 Data File : VD067901.D
 Acq On : 07 Dec 2020 15:58
 Operator : VA/SY
 Sample : L4967-01
 Misc : 6.07G/5.00ml/MSVOA D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-05-120720

Quant Time: Dec 08 03:51:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 02 08:05:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	236305	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	331352	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	287446	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	136694	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.33	65	115121	45.38	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	90.76%		
35) Dibromofluoromethane	7.92	113	98047	44.22	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	88.44%		
50) Toluene-d8	10.34	98	370027	46.36	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	92.72%		
62) 4-Bromofluorobenzene	12.64	95	118730	43.03	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	86.06%		

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD120720\
Data File : VD067901.D
Acq On : 07 Dec 2020 15:58
Operator : VA/SY
Sample : L4967-01
Misc : 6.07G/5.00ml/MSVOA D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-05-120720

Quant Time: Dec 08 03:51:32 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 02 08:05:58 2020
Response via : Initial Calibration

