

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD120919\
 Data File : VD064475.D
 Acq On : 09 Dec 2019 17:18
 Operator : VA/SY
 Sample : K6205-03
 Misc : 4.41G/5.00ml/MSVOA D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 K165-SB-3

Quant Time: Dec 10 02:58:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 29 07:18:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	693569	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	1038599	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	980436	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	491243	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	297165	47.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.68%	
35) Dibromofluoromethane	7.92	113	315728	48.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.50%	
50) Toluene-d8	10.34	98	1241105	50.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.70%	
62) 4-Bromofluorobenzene	12.64	95	426543	53.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.70%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD120919\
Data File : VD064475.D
Acq On : 09 Dec 2019 17:18
Operator : VA/SY
Sample : K6205-03
Misc : 4.41G/5.00ml/MSVOA D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
K165-SB-3

Quant Time: Dec 10 02:58:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D112719S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 29 07:18:03 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.98 min Scan# 1031

Delta R.T. 0.01 min

Lab File: VD064475.D

Acq: 09 Dec 2019 17:18

Instrument :

MSVOA_D

ClientSampleId :


K165-SB-3

Tot Ion:168 Resp: 693569

Ion Ratio Lower Upper




168 100

99 51.3 40.7 61.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VDC



#33

1,2-Dichloroethane-d4

Concen: 47.336 ug/l

RT: 8.33 min Scan# 1090

Delta R.T. -0.01 min

Lab File: VD064475.D


Acq: 09 Dec 2019 17:18

Tgt Ion: 65 Resp: 297165

Ion Ratio Lower Upper



65 100


67 51.9 0.0 102.0



Abundance Ion 64.90 (64.60 to 65.60): VDC

Ion 66.90 (66.60 to 67.60): VDC









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.87 min Scan# 1181
Delta R.T. 0.00 min
Lab File: VD064475.D
Acq: 09 Dec 2019 17:18

Instrument :
MSVOA_D
ClientSampleId :
K165-SB-3

Tot Ion:114 Resp: 1038599
Ion Ratio Lower Upper
114 100
63 19.4 0.0 40.0
88 15.2 0.0 29.2




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): V
Ion 87.90 (87.60 to 88.60): V





#35
Dibromofluoromethane
Concen: 48.249 ug/l
RT: 7.92 min Scan# 1020
Delta R.T. 0.00 min
Lab File: VD064475.D
Acq: 09 Dec 2019 17:18

Tgt Ion:113 Resp: 315728
Ion Ratio Lower Upper
113 100
111 102.2 79.4 119.2
192 22.9 18.6 28.0



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 50.350 ug/l

RT: 10.34 min Scan# 1431

Delta R.T. -0.01 min

Lab File: VD064475.D

Acq: 09 Dec 2019 17:18

Instrument :

MSVOA_D

ClientSampleId :





K165-SB-3

Tot Ion: 98 Resp: 1241105

Ion Ratio Lower Upper

98 100

100 65.0 52.6 79.0



#62

4-Bromofluorobenzene

Concen: 53.845 ug/l

RT: 12.64 min Scan# 1822

Delta R.T. 0.00 min

Lab File: VD064475.D

Acq: 09 Dec 2019 17:18




Tgt Ion: 95 Resp: 426543


Ion Ratio Lower Upper

95 100

174 96.3 0.0 197.2

176 93.2 0.0 189.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.65 min Scan# 1654
Delta R.T. 0.00 min
Lab File: VD064475.D
Acq: 09 Dec 2019 17:18

Instrument :
MSVOA_D
ClientSampleId :
K165-SB-3

Tot Ion:117 Resp: 980436
Ion Ratio Lower Upper
117 100
82 51.7 41.4 62.2
119 30.8 26.2 39.2




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.58 min Scan# 1983
Delta R.T. 0.00 min
Lab File: VD064475.D
Acq: 09 Dec 2019 17:18

Tgt Ion:152 Resp: 491243
Ion Ratio Lower Upper
152 100
115 54.4 26.7 80.1
150 156.2 0.0 347.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

