

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD120920\
 Data File : VD067918.D
 Acq On : 09 Dec 2020 10:33
 Operator : VA/SY
 Sample : VD1209SBL01
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1209SBL01

Quant Time: Dec 10 03:29:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 02 08:05:58 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	240919	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	329456	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	295569	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	144819	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.33	65	109873	42.48	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	84.96%		
35) Dibromofluoromethane	7.92	113	97786	44.36	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	88.72%		
50) Toluene-d8	10.34	98	369941	46.62	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	93.24%		
62) 4-Bromofluorobenzene	12.64	95	121923	44.44	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	88.88%		

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD120920\
Data File : VD067918.D
Acq On : 09 Dec 2020 10:33
Operator : VA/SY
Sample : VD1209SBL01
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1209SBL01

Quant Time: Dec 10 03:29:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120120S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 02 08:05:58 2020
Response via : Initial Calibration

