

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD121818\
 Data File : VD060627.D
 Acq On : 18 Dec 2018 18:03
 Operator : VA/AP
 Sample : J6445-03RE
 Misc : 5.06g/5ml/MSVOA D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EB02_1.5-3.5RE

Quant Time: Dec 19 03:25:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 06 01:49:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.40	168	34150	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	7.44	114	52877	50.00	ug/l	-0.02
63) Chlorobenzene-d5	11.27	117	53798	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.37	152	22483	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.78	65	26014	107.47	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	214.94%	
35) Dibromofluoromethane	6.32	113	23201	59.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.48%	
50) Toluene-d8	9.45	98	46543	41.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.82%	
62) 4-Bromofluorobenzene	12.43	95	15650	38.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.82%	
Target Compounds						
16) Acetone	3.16	43	8076	208.149	ug/l #	85
20) Methylene Chloride	3.74	84	7378	17.832	ug/l #	95
52) Toluene	9.55	92	2268	2.747	ug/l #	53

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD121818\
Data File : VD060627.D
Acq On : 18 Dec 2018 18:03
Operator : VA/AP
Sample : J6445-03RE
Misc : 5.06g/5ml/MSVOA D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EB02_1.5-3.5RE

Quant Time: Dec 19 03:25:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 06 01:49:20 2018
Response via : Initial Calibration

