


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD121818\
 Data File : VD060615.D
 Acq On : 18 Dec 2018 11:10
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
 Title : SW846 8260
 Last Update : Thu Dec 06 01:49:20 2018



AutoFind: Scans 1107, 1108, 1109; Background Corrected with Scan 1102

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.4	89882	PASS
75	95	30	60	40.1	186282	PASS
95	95	100	100	100.0	464192	PASS
96	95	5	9	6.4	29698	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	93.9	435840	PASS
175	174	5	9	6.9	30060	PASS
176	174	95	101	98.6	429717	PASS
177	176	5	9	6.4	27533	PASS