

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD122319\
 Data File : VD064560.D
 Acq On : 23 Dec 2019 11:10
 Operator : VA/SY
 Sample : VD1223SBL01
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1223SBL01

Quant Time: Dec 24 06:01:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 29 07:18:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	567255	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	834591	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	855449	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	437513	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	236819	46.12	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.24%	
35) Dibromofluoromethane	7.91	113	253452	48.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.40%	
50) Toluene-d8	10.34	98	1020217	51.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.02%	
62) 4-Bromofluorobenzene	12.64	95	379215	59.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	119.14%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD122319\
Data File : VD064560.D
Acq On : 23 Dec 2019 11:10
Operator : VA/SY
Sample : VD1223SBL01
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1223SBL01

Quant Time: Dec 24 06:01:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D112719S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 29 07:18:03 2019
Response via : Initial Calibration

