

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD122319\
 Data File : VD064577.D
 Acq On : 23 Dec 2019 19:24
 Operator : VA/SY
 Sample : K6118-10
 Misc : 8.8G/5.00ml/MSVOA D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-01-122319

Quant Time: Dec 24 06:30:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 29 07:18:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	409155	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	674358	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	616617	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	226147	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.34	65	190194	51.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.72%	
35) Dibromofluoromethane	7.92	113	214086	50.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.78%	
50) Toluene-d8	10.34	98	776546	48.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.04%	
62) 4-Bromofluorobenzene	12.64	95	250497	48.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.40%	
Target Compounds						
16) Acetone	4.17	43	7033	9.397	ug/l #	53
30) Chloroform	7.71	83	13740	1.942	ug/l	85
68) m/p-Xylenes	11.85	106	19340	2.217	ug/l	98
69) o-Xylene	12.18	106	12542	1.557	ug/l	98
80) 1,3,5-Trimethylbenzene	12.96	105	20407	1.535	ug/l	97
84) 1,2,4-Trimethylbenzene	13.28	105	65838	4.926	ug/l	98
95) Naphthalene	15.41	128	189200	22.722	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD122319\
Data File : VD064577.D
Acq On : 23 Dec 2019 19:24
Operator : VA/SY
Sample : K6118-10
Misc : 8.8G/5.00ml/MSVOA D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-01-122319

Quant Time: Dec 24 06:30:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D112719S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 29 07:18:03 2019
Response via : Initial Calibration

