

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122420\
 Data File : VD067950.D
 Acq On : 24 Dec 2020 17:05
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :

Quant Time: Dec 25 00:16:35 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 23 07:34:16 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	280226	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	470942	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.649	117	439202	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	211889	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	151262	49.99	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.98%
35) Dibromofluoromethane	7.920	113	151361	51.01	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.02%
50) Toluene-d8	10.344	98	559407	50.16	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.32%
62) 4-Bromofluorobenzene	12.632	95	191286	50.14	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.28%

Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.985	85	153006	53.61	ug/l	97
3) Chloromethane	2.209	50	172264	44.15	ug/l	99
4) Vinyl Chloride	2.350	62	169988	46.75	ug/l	92
5) Bromomethane	2.767	94	111316	48.29	ug/l	94
6) Chloroethane	2.920	64	103067	46.12	ug/l	94
7) Trichlorofluoromethane	3.273	101	284868	53.16	ug/l	98
8) Diethyl Ether	3.709	74	82059	53.41	ug/l	96
9) 1,1,2-Trichlorotrifluo...	4.091	101	167791	52.53	ug/l	99
10) Methyl Iodide	4.297	142	156946	48.30	ug/l	98
11) Tert butyl alcohol	5.214	59	52976	318.13	ug/l #	90
12) 1,1-Dichloroethene	4.067	96	156732	51.16	ug/l	95
13) Acrolein	3.920	56	49931	231.86	ug/l	97
14) Allyl chloride	4.714	41	260811	47.88	ug/l	96
15) Acrylonitrile	5.426	53	191027	259.71	ug/l	99
16) Acetone	4.156	43	154376	267.21	ug/l	95
17) Carbon Disulfide	4.409	76	532571	49.88	ug/l	98
18) Methyl Acetate	4.720	43	85297	49.96	ug/l	98
19) Methyl tert-butyl Ether	5.479	73	348923	54.77	ug/l	95
20) Methylene Chloride	4.961	84	191511	51.47	ug/l	98
21) trans-1,2-Dichloroethene	5.473	96	183064	50.00	ug/l	97
22) Diisopropyl ether	6.361	45	562842	51.30	ug/l	96
23) Vinyl Acetate	6.303	43	1483930	266.08	ug/l	100
24) 1,1-Dichloroethane	6.261	63	321960	48.48	ug/l	96
25) 2-Butanone	7.214	43	223307	262.09	ug/l	97
26) 2,2-Dichloropropane	7.208	77	270672	51.72	ug/l	100
27) cis-1,2-Dichloroethene	7.208	96	200232	51.19	ug/l	98
28) Bromochloromethane	7.550	49	124343	49.66	ug/l	95
29) Tetrahydrofuran	7.561	42	152958	265.40	ug/l	96
30) Chloroform	7.708	83	332194	51.06	ug/l	96
31) Cyclohexane	7.985	56	299756	48.46	ug/l	96
32) 1,1,1-Trichloroethane	7.908	97	292136	52.42	ug/l	98
36) 1,1-Dichloropropene	8.114	75	264080	52.76	ug/l	100
37) Ethyl Acetate	7.297	43	109449	54.04	ug/l	98
38) Carbon Tetrachloride	8.102	117	252737	52.86	ug/l	96
39) Methylcyclohexane	9.361	83	307409	53.56	ug/l	94
40) Benzene	8.355	78	743608	51.21	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122420\
 Data File : VD067950.D
 Acq On : 24 Dec 2020 17:05
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :

Quant Time: Dec 25 00:16:35 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
 Quant Title : SW846 8260
 QLast Update : Wed Dec 23 07:34:16 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.520	41	59843	52.49	ug/l	93
42) 1,2-Dichloroethane	8.426	62	205303	52.43	ug/l	98
43) Isopropyl Acetate	8.455	43	197373	53.09	ug/l	98
44) Trichloroethene	9.114	130	189176	51.97	ug/l	100
45) 1,2-Dichloropropane	9.391	63	190709	51.59	ug/l #	89
46) Dibromomethane	9.479	93	101779	56.65	ug/l	95
47) Bromodichloromethane	9.667	83	258059	53.23	ug/l	98
48) Methyl methacrylate	9.461	41	106619	55.49	ug/l	96
49) 1,4-Dioxane	9.467	88	21804	1111.36	ug/l	93
51) 4-Methyl-2-Pentanone	10.232	43	537499	279.21	ug/l	98
52) Toluene	10.402	92	468487	52.27	ug/l	100
53) t-1,3-Dichloropropene	10.626	75	240036	54.75	ug/l	94
54) cis-1,3-Dichloropropene	10.091	75	285851	52.69	ug/l	99
55) 1,1,2-Trichloroethane	10.802	97	133323	53.09	ug/l	95
56) Ethyl methacrylate	10.661	69	163957	58.11	ug/l	96
57) 1,3-Dichloropropane	10.949	76	236291	53.51	ug/l	97
58) 2-Chloroethyl Vinyl ether	9.944	63	424413	298.91	ug/l	98
59) 2-Hexanone	10.985	43	368980	291.67	ug/l	99
60) Dibromochloromethane	11.143	129	173182	55.67	ug/l	99
61) 1,2-Dibromoethane	11.249	107	130346	54.10	ug/l	100
64) Tetrachloroethene	10.879	164	155329	50.41	ug/l	97
65) Chlorobenzene	11.673	112	482809	50.96	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.743	131	177854	52.59	ug/l	98
67) Ethyl Benzene	11.743	91	882278	52.21	ug/l	99
68) m/p-Xylenes	11.855	106	679434	105.51	ug/l	100
69) o-Xylene	12.179	106	309381	53.69	ug/l	98
70) Styrene	12.196	104	534418	53.67	ug/l	99
71) Bromoform	12.361	173	93726	54.83	ug/l #	99
73) Isopropylbenzene	12.479	105	825945	52.51	ug/l	100
74) N-amyl acetate	12.290	43	188273	52.73	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.732	83	154180	52.84	ug/l	99
76) 1,2,3-Trichloropropane	12.785	75	197666	110.80	ug/l #	100
77) Bromobenzene	12.761	156	189544	52.70	ug/l	99
78) n-propylbenzene	12.820	91	1034367	52.91	ug/l	99
79) 2-Chlorotoluene	12.908	91	567794	51.40	ug/l	100
80) 1,3,5-Trimethylbenzene	12.961	105	693399	53.09	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.526	75	46984	53.99	ug/l	97
82) 4-Chlorotoluene	13.002	91	606755	52.27	ug/l	98
83) tert-Butylbenzene	13.226	119	579186	53.91	ug/l	100
84) 1,2,4-Trimethylbenzene	13.267	105	696614	53.18	ug/l	98
85) sec-Butylbenzene	13.402	105	836636	53.22	ug/l	100
86) p-Isopropyltoluene	13.514	119	751375	54.32	ug/l	99
87) 1,3-Dichlorobenzene	13.514	146	367559	51.39	ug/l	100
88) 1,4-Dichlorobenzene	13.596	146	358455	50.74	ug/l	99
89) n-Butylbenzene	13.843	91	726309	52.93	ug/l	99
90) Hexachloroethane	14.108	117	144887	51.26	ug/l	99
91) 1,2-Dichlorobenzene	13.890	146	318438	52.88	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.508	75	23125	50.63	ug/l	99
93) 1,2,4-Trichlorobenzene	15.161	180	192861	53.94	ug/l	98
94) Hexachlorobutadiene	15.267	225	106359	50.83	ug/l	97
95) Naphthalene	15.402	128	362727	57.40	ug/l	98
96) 1,2,3-Trichlorobenzene	15.590	180	166682	53.96	ug/l	96


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122420\
Data File : VD067950.D
Acq On : 24 Dec 2020 17:05
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :

Quant Time: Dec 25 00:16:35 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
Quant Title : SW846 8260
QLast Update : Wed Dec 23 07:34:16 2020
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.00 ug/l
RT: 7.985 min Scan# 1031
Delta R.T. -0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:168 Resp: 280226
Ion Ratio Lower Upper
168 100
99 59.2 48.0 72.0



#2
Dichlorodifluoromethane
Concen: 53.61 ug/l
RT: 1.985 min Scan# 11
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 85 Resp: 153006
Ion Ratio Lower Upper
85 100
87 32.4 17.1 51.3









#3
Chloromethane
Concen: 44.15 ug/l
RT: 2.209 min Scan# 49
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 50 Resp: 172264
Ion Ratio Lower Upper
50 100
52 33.5 26.2 39.4



#4
Vinyl Chloride
Concen: 46.75 ug/l
RT: 2.350 min Scan# 73
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 62 Resp: 169988
Ion Ratio Lower Upper
62 100
64 30.5 28.1 42.1









#5
Bromomethane
Concen: 48.29 ug/l
RT: 2.767 min Scan# 144
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 94 Resp: 111316
Ion Ratio Lower Upper
94 100
96 103.3 78.0 117.0



#6
Chloroethane
Concen: 46.12 ug/l
RT: 2.920 min Scan# 170
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 64 Resp: 103067
Ion Ratio Lower Upper
64 100
66 33.4 24.1 36.1









#7
Trichlorofluoromethane
Concen: 53.16 ug/l
RT: 3.273 min Scan# 230
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:101 Resp: 284868
Ion Ratio Lower Upper
101 100
103 64.9 50.4 75.6



#8
Diethyl Ether
Concen: 53.41 ug/l
RT: 3.709 min Scan# 304
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 74 Resp: 82059
Ion Ratio Lower Upper
74 100
45 99.5 51.8 155.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 52.53 ug/l
RT: 4.091 min Scan# 369
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:101 Resp: 167791
Ion Ratio Lower Upper
101 100
85 44.4 34.2 51.2
151 71.5 57.6 86.4



#10
Methyl Iodide
Concen: 48.30 ug/l
RT: 4.297 min Scan# 404
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion:142 Resp: 156946
Ion Ratio Lower Upper
142 100
127 44.5 34.4 51.6
141 14.8 11.8 17.8









#11
Tert butyl alcohol
Concen: 318.13 ug/l
RT: 5.214 min Scan# 560
Delta R.T. -0.012 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 59 Resp: 52976
Ion Ratio Lower Upper
59 100
57 5.5 7.4 11.2#



#12
1,1-Dichloroethene
Concen: 51.16 ug/l
RT: 4.067 min Scan# 365
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 96 Resp: 156732
Ion Ratio Lower Upper
96 100
61 160.0 133.6 200.4
98 66.0 54.8 82.2









#13
Acrolein
Concen: 231.86 ug/l
RT: 3.920 min Scan# 340
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 56 Resp: 49931
Ion Ratio Lower Upper
56 100
55 70.6 58.5 87.7



#14
Allyl chloride
Concen: 47.88 ug/l
RT: 4.714 min Scan# 475
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05


Tgt Ion: 41 Resp: 260811
Ion Ratio Lower Upper
41 100
39 65.9 50.5 75.7
76 37.2 27.0 40.4








#15
Acrylonitrile
Concen: 259.71 ug/l
RT: 5.426 min Scan# 596
Delta R.T. 0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Instrument :
MSVOA_D
ClientSampled :







Tgt Ion: 53 Resp: 191027
Ion Ratio Lower Upper
53 100
52 80.9 65.3 97.9
51 37.0 29.4 44.2



#16
Acetone
Concen: 267.21 ug/l
RT: 4.156 min Scan# 380
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05


Tgt Ion: 43 Resp: 154376
Ion Ratio Lower Upper
43 100
58 30.4 22.3 33.5








#17
Carbon Disulfide
Concen: 49.88 ug/l
RT: 4.409 min Scan# 423
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05


Instrument :
MSVOA_D
ClientSampleId :





Tgt Ion: 76 Resp: 532571
Ion Ratio Lower Upper
76 100
78 8.0 7.1 10.7




#18
Methyl Acetate
Concen: 49.96 ug/l
RT: 4.720 min Scan# 476
Delta R.T. 0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05



Tgt Ion: 43 Resp: 85297
Ion Ratio Lower Upper
43 100
74 23.6 18.2 27.2









#19
Methyl tert-butyl Ether
Concen: 54.77 ug/l
RT: 5.479 min Scan# 605
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 73 Resp: 348923
Ion Ratio Lower Upper
73 100
57 22.1 19.8 29.6



#20
Methylene Chloride
Concen: 51.47 ug/l
RT: 4.961 min Scan# 517
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05


Tgt Ion: 84 Resp: 191511
Ion Ratio Lower Upper
84 100
49 122.3 98.3 147.5
51 34.9 30.3 45.5
86 62.9 49.0 73.4








#21
trans-1,2-Dichloroethene
Concen: 50.00 ug/l
RT: 5.473 min Scan# 604
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05


Instrument :
MSVOA_D
ClientSampleId :





Tgt Ion: 96 Resp: 183064
Ion Ratio Lower Upper
96 100
61 134.4 109.9 164.9
98 63.5 49.0 73.4




#22
Diisopropyl ether
Concen: 51.30 ug/l
RT: 6.361 min Scan# 755
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05



Tgt Ion: 45 Resp: 562842
Ion Ratio Lower Upper
45 100
43 54.1 40.2 60.4
87 25.2 19.2 28.8
59 10.6 8.7 13.1









#23
 Vinyl Acetate
 Concen: 266.08 ug/l
 RT: 6.303 min Scan# 745
 Delta R.T. -0.006 min
 Lab File: VD067950.D
 Acq: 24 Dec 2020 17:05




Instrument :
 MSVOA_D
 ClientSampleId :


Tgt Ion: 43 Resp: 1483930
 Ion Ratio Lower Upper
 43 100
 86 10.0 7.9 11.9



#24
 1,1-Dichloroethane
 Concen: 48.48 ug/l
 RT: 6.261 min Scan# 738
 Delta R.T. -0.006 min
 Lab File: VD067950.D
 Acq: 24 Dec 2020 17:05


Tgt Ion: 63 Resp: 321960
 Ion Ratio Lower Upper
 63 100
 98 7.7 2.9 8.6
 100 3.9 1.9 5.7








#25
2-Butanone
Concen: 262.09 ug/l
RT: 7.214 min Scan# 900
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05


Instrument :
MSVOA_D
ClientSampled :





Tgt Ion: 43 Resp: 223307
Ion Ratio Lower Upper
43 100
72 27.7 20.8 31.2




#26
2,2-Dichloropropane
Concen: 51.72 ug/l
RT: 7.208 min Scan# 899
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05



Tgt Ion: 77 Resp: 270672
Ion Ratio Lower Upper
77 100
97 22.9 11.4 34.2





#27

cis-1,2-Dichloroethene

Concen: 51.19 ug/l

RT: 7.208 min Scan# 899





Delta R.T. -0.006 min

Lab File: VD067950.D

Acq: 24 Dec 2020 17:05

Tgt Ion:	96	Resp:	200232
Ion	Ratio	Lower	Upper
96	100		
61	146.7	0.0	295.4
98	62.4	0.0	130.4

Instrument :
MSVOA_D
ClientSampleId :



#28

Bromochloromethane

Concen: 49.66 ug/l




RT: 7.550 min Scan# 957


Delta R.T. 0.000 min

Lab File: VD067950.D

Acq: 24 Dec 2020 17:05

Tgt Ion:	49	Resp:	124343
Ion	Ratio	Lower	Upper
49	100		
129	1.2	0.0	4.4
130	73.0	54.8	82.2





#29

Tetrahydrofuran

Concen: 265.40 ug/l

RT: 7.561 min Scan# 959





Delta R.T. -0.006 min

Lab File: VD067950.D

Acq: 24 Dec 2020 17:05

Instrument :
MSVOA_D
ClientSampleId :

Tgt Ion:	42	Resp:	152958
Ion	Ratio	Lower	Upper
42	100		
72	43.3	32.9	49.3
71	39.5	29.9	44.9



#30

Chloroform

Concen: 51.06 ug/l




RT: 7.708 min Scan# 984


Delta R.T. -0.006 min

Lab File: VD067950.D

Acq: 24 Dec 2020 17:05


Tgt Ion:	83	Resp:	332194
Ion	Ratio	Lower	Upper
83	100		
85	60.1	50.5	75.7








#31
Cyclohexane
Concen: 48.46 ug/l
RT: 7.985 min Scan# 1031
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Instrument :
MSVOA_D
ClientSampled :







Tgt Ion: 56 Resp: 299756
Ion Ratio Lower Upper
56 100
69 31.5 22.7 34.1
84 83.1 64.2 96.4



#32
1,1,1-Trichloroethane
Concen: 52.42 ug/l
RT: 7.908 min Scan# 1018
Delta R.T. 0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 97 Resp: 292136
Ion Ratio Lower Upper
97 100
99 62.7 51.0 76.6
61 42.8 35.8 53.8









#33
1,2-Dichloroethane-d4
Concen: 49.99 ug/l
RT: 8.332 min Scan# 1090
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 65 Resp: 151262
Ion Ratio Lower Upper
65 100
67 53.3 0.0 110.0



#34
1,4-Difluorobenzene
Concen: 50.00 ug/l
RT: 8.867 min Scan# 1181
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 114 Resp: 470942
Ion Ratio Lower Upper
114 100
63 22.4 0.0 41.4
88 16.6 0.0 33.2





#35

Dibromofluoromethane

Concen: 51.01 ug/l

RT: 7.920 min Scan# 1020

Delta R.T. 0.000 min

Lab File: VD067950.D

Acq: 24 Dec 2020 17:05





Tgt Ion:113 Resp: 151361

Ion Ratio Lower Upper

113 100

111 102.4 79.3 118.9

192 17.9 14.5 21.7



#36

1,1-Dichloropropene

Concen: 52.76 ug/l

RT: 8.114 min Scan# 1053

Delta R.T. 0.000 min

Lab File: VD067950.D

Acq: 24 Dec 2020 17:05




Tgt Ion: 75 Resp: 264080


Ion Ratio Lower Upper

75 100

110 34.0 17.0 51.0

77 30.8 24.6 37.0





#37

Ethyl Acetate

Concen: 54.04 ug/l

RT: 7.297 min Scan# 914





Delta R.T. 0.000 min

Lab File: VD067950.D

Acq: 24 Dec 2020 17:05

Tgt Ion: 43 Resp: 109449

Ion	Ratio	Lower	Upper
43	100		
61	13.9	10.2	15.4
70	9.6	8.3	12.5



#38

Carbon Tetrachloride

Concen: 52.86 ug/l

RT: 8.102 min Scan# 1051




Delta R.T. 0.000 min


Lab File: VD067950.D

Acq: 24 Dec 2020 17:05

Tgt Ion: 117 Resp: 252737

Ion	Ratio	Lower	Upper
117	100		
119	90.6	76.2	114.4
121	31.1	25.9	38.9





#39

Methylcyclohexane

Concen: 53.56 ug/l

RT: 9.361 min Scan# 1265





Delta R.T. 0.006 min

Lab File: VD067950.D

Acq: 24 Dec 2020 17:05

Tgt Ion: 83 Resp: 307409

Ion	Ratio	Lower	Upper
83	100		
55	72.8	63.5	95.3
98	48.3	37.3	55.9



#40

Benzene

Concen: 51.21 ug/l

RT: 8.355 min Scan# 1094




Delta R.T. 0.000 min


Lab File: VD067950.D

Acq: 24 Dec 2020 17:05

Tgt Ion: 78 Resp: 743608


Ion	Ratio	Lower	Upper
78	100		
77	23.2	20.1	30.1








#41
Methacrylonitrile
Concen: 52.49 ug/l
RT: 7.520 min Scan# 952
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion:	41	Resp:	59843
Ion Ratio	Lower	Upper	
41	100		
39	53.0	46.2	69.2
67	77.4	55.1	82.7
52	30.2	23.5	35.3



#42
1,2-Dichloroethane
Concen: 52.43 ug/l
RT: 8.426 min Scan# 1106
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05


Tgt Ion:	62	Resp:	205303
Ion Ratio	Lower	Upper	
62	100		
98	9.6	0.0	20.4








#43
Isopropyl Acetate
Concen: 53.09 ug/l
RT: 8.455 min Scan# 1111
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Instrument :
MSVOA_D
ClientSampleId :







Tgt Ion: 43 Resp: 197373
Ion Ratio Lower Upper
43 100
61 31.0 24.5 36.7
87 13.7 9.8 14.8



#44
Trichloroethene
Concen: 51.97 ug/l
RT: 9.114 min Scan# 1223
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion:130 Resp: 189176
Ion Ratio Lower Upper
130 100
95 100.4 0.0 200.8









#45

1,2-Dichloropropane
Concen: 51.59 ug/l
RT: 9.391 min Scan# 1270
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Instrument :
MSVOA_D
ClientSampled :




Tgt Ion: 63 Resp: 190709
Ion Ratio Lower Upper
63 100
65 34.2 22.7 34.1#




#46

Dibromomethane
Concen: 56.65 ug/l
RT: 9.479 min Scan# 1285
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 93 Resp: 101779
Ion Ratio Lower Upper
93 100
95 78.6 68.0 102.0
174 89.9 74.7 112.1









#47
Bromodichloromethane
Concen: 53.23 ug/l
RT: 9.667 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 83 Resp: 258059
Ion Ratio Lower Upper
83 100
85 64.2 50.5 75.7
127 8.1 7.2 10.8



#48
Methyl methacrylate
Concen: 55.49 ug/l
RT: 9.461 min Scan# 1282
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 41 Resp: 106619
Ion Ratio Lower Upper
41 100
69 83.4 63.9 95.9
39 49.1 37.4 56.0





#49

1,4-Dioxane

Concen: 1111.36 ug/l

RT: 9.467 min Scan# 1283

Delta R.T. 0.000 min

Lab File: VD067950.D

Acq: 24 Dec 2020 17:05





Tgt Ion: 88 Resp: 21804

Ion Ratio Lower Upper

88 100

43 40.9 27.3 40.9

58 72.8 54.8 82.2



#50

Toluene-d8

Concen: 50.16 ug/l

RT: 10.344 min Scan# 1432

Delta R.T. 0.000 min

Lab File: VD067950.D




Acq: 24 Dec 2020 17:05


Tgt Ion: 98 Resp: 559407

Ion Ratio Lower Upper

98 100

100 63.8 52.5 78.7









#51
4-Methyl-2-Pentanone
Concen: 279.21 ug/l
RT: 10.232 min Scan# 1413
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 43 Resp: 537499
Ion Ratio Lower Upper
43 100
58 38.8 30.0 45.0



#52
Toluene
Concen: 52.27 ug/l
RT: 10.402 min Scan# 1442
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 92 Resp: 468487
Ion Ratio Lower Upper
92 100
91 173.6 138.5 207.7









#53
t-1,3-Dichloropropene
Concen: 54.75 ug/l
RT: 10.626 min Scan# 1480
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 75 Resp: 240036
Ion Ratio Lower Upper
75 100
77 29.8 26.8 40.2



#54
cis-1,3-Dichloropropene
Concen: 52.69 ug/l
RT: 10.091 min Scan# 1389
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 75 Resp: 285851
Ion Ratio Lower Upper
75 100
77 30.8 25.3 37.9
39 45.2 35.7 53.5









#55
1,1,2-Trichloroethane
Concen: 53.09 ug/l
RT: 10.802 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:	97	Resp:	133323
Ion	Ratio	Lower	Upper
97	100		
83	88.8	66.4	99.6
85	57.7	43.4	65.0
99	64.9	49.4	74.0



#56
Ethyl methacrylate
Concen: 58.11 ug/l
RT: 10.661 min Scan# 1486
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion:	69	Resp:	163957
Ion	Ratio	Lower	Upper
69	100		
41	66.7	56.4	84.6
39	29.5	24.6	36.8









#57
1,3-Dichloropropane
Concen: 53.51 ug/l
RT: 10.949 min Scan# 1535
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 76 Resp: 236291
Ion Ratio Lower Upper
76 100
78 33.1 25.2 37.8



#58
2-Chloroethyl Vinyl ether
Concen: 298.91 ug/l
RT: 9.944 min Scan# 1364
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 63 Resp: 424413
Ion Ratio Lower Upper
63 100
106 27.8 21.6 32.4









#59
2-Hexanone
Concen: 291.67 ug/l
RT: 10.985 min Scan# 1541
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 43 Resp: 368980
Ion Ratio Lower Upper
43 100
58 54.1 27.5 82.4



#60
Dibromochloromethane
Concen: 55.67 ug/l
RT: 11.143 min Scan# 1568
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 129 Resp: 173182
Ion Ratio Lower Upper
129 100
127 76.9 38.8 116.4









#61
1,2-Dibromoethane
Concen: 54.10 ug/l
RT: 11.249 min Scan# 1586
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:107 Resp: 130346
Ion Ratio Lower Upper
107 100
109 94.2 75.4 113.2



#62
4-Bromofluorobenzene
Concen: 50.14 ug/l
RT: 12.632 min Scan# 1821
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 95 Resp: 191286
Ion Ratio Lower Upper
95 100
174 77.0 0.0 150.6
176 74.9 0.0 151.2









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 11.649 min Scan# 1654
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion:117 Resp: 439202
Ion Ratio Lower Upper
117 100
82 56.9 45.4 68.0
119 31.9 26.7 40.1



#64
Tetrachloroethene
Concen: 50.41 ug/l
RT: 10.879 min Scan# 1523
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion:164 Resp: 155329
Ion Ratio Lower Upper
164 100
166 128.8 99.4 149.2
129 98.9 79.0 118.4
131 93.2 71.0 106.4









#65
Chlorobenzene
Concen: 50.96 ug/l
RT: 11.673 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:112 Resp: 482809
Ion Ratio Lower Upper
112 100
114 33.3 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 52.59 ug/l
RT: 11.743 min Scan# 1670
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion:131 Resp: 177854
Ion Ratio Lower Upper
131 100
133 93.4 48.4 145.0
119 66.5 33.4 100.2









#67
Ethyl Benzene
Concen: 52.21 ug/l
RT: 11.743 min Scan# 1670
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 91 Resp: 882278
Ion Ratio Lower Upper
91 100
106 31.6 24.6 37.0



#68
m/p-Xylenes
Concen: 105.51 ug/l
RT: 11.855 min Scan# 1689
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion:106 Resp: 679434
Ion Ratio Lower Upper
106 100
91 199.5 159.9 239.9









#69
o-Xylene
Concen: 53.69 ug/l
RT: 12.179 min Scan# 1744
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:106 Resp: 309381
Ion Ratio Lower Upper
106 100
91 208.8 106.0 317.9



#70
Styrene
Concen: 53.67 ug/l
RT: 12.196 min Scan# 1747
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion:104 Resp: 534418
Ion Ratio Lower Upper
104 100
78 49.4 38.6 58.0
103 54.7 44.2 66.4









#71
Bromoform
Concen: 54.83 ug/l
RT: 12.361 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:173 Resp: 93726
Ion Ratio Lower Upper
173 100
175 48.8 24.6 73.8
254 0.1 0.0 0.0#



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 13.573 min Scan# 1981
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion:152 Resp: 211889
Ion Ratio Lower Upper
152 100
115 59.2 29.3 88.0
150 174.0 0.0 340.0









#73
Isopropylbenzene
Concen: 52.51 ug/l
RT: 12.479 min Scan# 1795
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:105 Resp: 825945
Ion Ratio Lower Upper
105 100
120 26.7 13.4 40.1



#74
N-ethyl acetate
Concen: 52.73 ug/l
RT: 12.290 min Scan# 1763
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 43 Resp: 188273
Ion Ratio Lower Upper
43 100
70 44.5 34.3 51.5
55 26.9 21.2 31.8
61 25.2 18.4 27.6









#75
1,1,2,2-Tetrachloroethane
Concen: 52.84 ug/l
RT: 12.732 min Scan# 1838
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 83 Resp: 154180
Ion Ratio Lower Upper
83 100
131 10.2 5.0 14.9
85 64.7 32.1 96.3



#76
1,2,3-Trichloropropane
Concen: 110.80 ug/l
RT: 12.785 min Scan# 1847
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 75 Resp: 197666
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 52.70 ug/l
RT: 12.761 min Scan# 1843
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:156 Resp: 189544
Ion Ratio Lower Upper
156 100
77 191.5 95.9 287.6
158 94.7 46.8 140.3



#78
n-propylbenzene
Concen: 52.91 ug/l
RT: 12.820 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 91 Resp: 1034367
Ion Ratio Lower Upper
91 100
120 22.6 11.2 33.5









#79
2-Chlorotoluene
Concen: 51.40 ug/l
RT: 12.908 min Scan# 1868
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampled :


Tgt Ion: 91 Resp: 567794
Ion Ratio Lower Upper
91 100
126 33.5 16.8 50.3



#80
1,3,5-Trimethylbenzene
Concen: 53.09 ug/l
RT: 12.961 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 105 Resp: 693399
Ion Ratio Lower Upper
105 100
120 50.7 24.9 74.6









#81
trans-1,4-Dichloro-2-butene
Concen: 53.99 ug/l
RT: 12.526 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 75 Resp: 46984
Ion Ratio Lower Upper
75 100
53 90.7 74.9 112.3
89 45.0 33.8 50.8



#82
4-Chlorotoluene
Concen: 52.27 ug/l
RT: 13.002 min Scan# 1884
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 91 Resp: 606755
Ion Ratio Lower Upper
91 100
126 32.9 17.0 51.0









#83
tert-Butylbenzene
Concen: 53.91 ug/l
RT: 13.226 min Scan# 1922
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:119 Resp: 579186
Ion Ratio Lower Upper
119 100
91 65.1 32.7 98.1
134 26.7 13.3 39.8



#84
1,2,4-Trimethylbenzene
Concen: 53.18 ug/l
RT: 13.267 min Scan# 1929
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion:105 Resp: 696614
Ion Ratio Lower Upper
105 100
120 46.4 22.7 68.0









#85
 sec-Butylbenzene
 Concen: 53.22 ug/l
 RT: 13.402 min Scan# 1952
 Delta R.T. 0.000 min
 Lab File: VD067950.D
 Acq: 24 Dec 2020 17:05




Instrument :
 MSVOA_D
 ClientSampleId :


Tgt Ion:105 Resp: 836636
 Ion Ratio Lower Upper
 105 100
 134 19.5 9.7 29.1



#86
 p-Isopropyltoluene
 Concen: 54.32 ug/l
 RT: 13.514 min Scan# 1971
 Delta R.T. -0.006 min
 Lab File: VD067950.D
 Acq: 24 Dec 2020 17:05

Tgt Ion:119 Resp: 751375
 Ion Ratio Lower Upper
 119 100
 134 25.9 13.3 39.9
 91 23.2 11.4 34.2









#87
1,3-Dichlorobenzene
Concen: 51.39 ug/l
RT: 13.514 min Scan# 1971
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:146 Resp: 367559
Ion Ratio Lower Upper
146 100
111 40.5 20.3 60.9
148 63.9 31.8 95.4



#88
1,4-Dichlorobenzene
Concen: 50.74 ug/l
RT: 13.596 min Scan# 1985
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion:146 Resp: 358455
Ion Ratio Lower Upper
146 100
111 39.2 19.4 58.2
148 63.3 31.9 95.9









#89
n-Butylbenzene
Concen: 52.93 ug/l
RT: 13.843 min Scan# 2027
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion: 91 Resp: 726309
Ion Ratio Lower Upper
91 100
92 54.1 26.8 80.3
134 25.0 12.3 36.9



#90
Hexachloroethane
Concen: 51.26 ug/l
RT: 14.108 min Scan# 2072
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 117 Resp: 144887
Ion Ratio Lower Upper
117 100
201 68.0 34.3 102.8









#91
1,2-Dichlorobenzene
Concen: 52.88 ug/l
RT: 13.890 min Scan# 2035
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:146 Resp: 318438
Ion Ratio Lower Upper
146 100
111 41.7 20.8 62.5
148 63.1 31.6 95.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 50.63 ug/l
RT: 14.508 min Scan# 2140
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion: 75 Resp: 23125
Ion Ratio Lower Upper
75 100
155 77.7 39.8 119.3
157 101.6 50.6 151.9









#93
1,2,4-Trichlorobenzene
Concen: 53.94 ug/l
RT: 15.161 min Scan# 2251
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05




Instrument :
MSVOA_D
ClientSampleId :


Tgt Ion:180 Resp: 192861
Ion Ratio Lower Upper
180 100
182 94.1 48.0 144.2
145 31.2 15.8 47.3



#94
Hexachlorobutadiene
Concen: 50.83 ug/l
RT: 15.267 min Scan# 2269
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion:225 Resp: 106359
Ion Ratio Lower Upper
225 100
223 64.4 33.4 100.2
227 65.1 31.7 95.1









#95
Naphthalene
Concen: 57.40 ug/l
RT: 15.402 min Scan# 2292
Delta R.T. 0.000 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Instrument :
MSVOA_D
ClientSampleId :

Tgt Ion:128 Resp: 362727
Ion Ratio Lower Upper
128 100
127 12.9 11.2 16.8
129 10.9 8.6 12.8



#96
1,2,3-Trichlorobenzene
Concen: 53.96 ug/l
RT: 15.590 min Scan# 2324
Delta R.T. -0.006 min
Lab File: VD067950.D
Acq: 24 Dec 2020 17:05

Tgt Ion:180 Resp: 166682
Ion Ratio Lower Upper
180 100
182 95.6 50.2 150.6
145 33.4 17.0 51.0

