

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122420\
 Data File : VD067950.D
 Acq On : 24 Dec 2020 17:05
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 25 00:16:35 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
 Quant Title : Sw846 8260
 QLast Update : Wed Dec 23 07:34:16 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.509	0.546	-7.3	95	0.00
3 P	Chloromethane	0.696	0.615	11.6	85	0.00
4 C	Vinyl Chloride	0.649	0.607	6.5#	87	0.00
5 T	Bromomethane	0.467	0.397	15.0	85	0.00
6 T	Chloroethane	0.399	0.368	7.8	88	0.00
7 T	Trichlorofluoromethane	0.956	1.017	-6.4	100	0.00
8 T	Diethyl Ether	0.274	0.293	-6.9	96	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.570	0.599	-5.1	99	0.00
10 T	Methyl Iodide	0.498	0.560	-12.4	87	0.00
11 T	Tert butyl alcohol	0.046	0.038	17.4	109	-0.01
12 CM	1,1-Dichloroethene	0.547	0.559	-2.2#	94	0.00
13 T	Acrolein	0.038	0.036	5.3	84	0.00
14 T	Allyl chloride	0.972	0.931	4.2	85	0.00
15 T	Acrylonitrile	0.131	0.136	-3.8	96	0.00
16 T	Acetone	0.103	0.110	-6.8	105	0.00
17 T	Carbon Disulfide	1.905	1.901	0.2	89	0.00
18 T	Methyl Acetate	0.305	0.304	0.3	93	0.00
19 T	Methyl tert-butyl Ether	1.137	1.245	-9.5	98	0.00
20 T	Methylene Chloride	0.849	0.683	19.6	88	0.00
21 T	trans-1,2-Dichloroethene	0.653	0.653	0.0	89	0.00
22 T	Diisopropyl ether	1.958	2.009	-2.6	90	0.00
23 T	Vinyl Acetate	0.995	1.059	-6.4	91	0.00
24 P	1,1-Dichloroethane	1.185	1.149	3.0	90	0.00
25 T	2-Butanone	0.152	0.159	-4.6	97	0.00
26 T	2,2-Dichloropropane	0.934	0.966	-3.4	97	0.00
27 T	cis-1,2-Dichloroethene	0.698	0.715	-2.4	92	0.00
28 T	Bromochloromethane	0.447	0.444	0.7	87	0.00
29 T	Tetrahydrofuran	0.103	0.109	-5.8	96	0.00
30 C	Chloroform	1.161	1.185	-2.1#	92	0.00
31 T	Cyclohexane	1.104	1.070	3.1	93	0.00
32 T	1,1,1-Trichloroethane	0.994	1.043	-4.9	98	0.00
33 S	1,2-Dichloroethane-d4	0.540	0.540	0.0	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.315	0.321	-1.9	92	0.00
36 T	1,1-Dichloropropene	0.531	0.561	-5.6	93	0.00
37 T	Ethyl Acetate	0.215	0.232	-7.9	96	0.00
38 T	Carbon Tetrachloride	0.508	0.537	-5.7	98	0.00
39 T	Methylcyclohexane	0.609	0.653	-7.2	98	0.00
40 TM	Benzene	1.542	1.579	-2.4	90	0.00
41 T	Methacrylonitrile	0.121	0.127	-5.0	93	0.00
42 TM	1,2-Dichloroethane	0.416	0.436	-4.8	96	0.00
43 T	Isopropyl Acetate	0.395	0.419	-6.1	93	0.00
44 TM	Trichloroethene	0.386	0.402	-4.1	95	0.00
45 C	1,2-Dichloropropane	0.392	0.405	-3.3#	93	0.00
46 T	Dibromomethane	0.191	0.216	-13.1	103	0.00
47 T	Bromodichloromethane	0.515	0.548	-6.4	95	0.00
48 T	Methyl methacrylate	0.204	0.226	-10.8	96	0.00
49 T	1,4-Dioxane	0.002	0.002	0.0	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122420\
 Data File : VD067950.D
 Acq On : 24 Dec 2020 17:05
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 25 00:16:35 2020
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
 Quant Title : Sw846 8260
 QLast Update : Wed Dec 23 07:34:16 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50 S	Toluene-d8	1.184	1.188	-0.3	93	0.00
51 T	4-Methyl-2-Pentanone	0.204	0.228	-11.8	97	0.00
52 CM	Toluene	0.952	0.995	-4.5#	92	0.00
53 T	t-1,3-Dichloropropene	0.466	0.510	-9.4	96	0.00
54 T	cis-1,3-Dichloropropene	0.576	0.607	-5.4	94	0.00
55 T	1,1,2-Trichloroethane	0.267	0.283	-6.0	97	0.00
56 T	Ethyl methacrylate	0.300	0.348	-16.0	99	0.00
57 T	1,3-Dichloropropane	0.469	0.502	-7.0	95	0.00
58 T	2-Chloroethyl vinyl ether	0.151	0.180	-19.2	95	0.00
59 T	2-Hexanone	0.134	0.157	-17.2	101	0.00
60 T	Dibromochloromethane	0.330	0.368	-11.5	101	0.00
61 T	1,2-Dibromoethane	0.256	0.277	-8.2	98	0.00
62 S	4-Bromofluorobenzene	0.405	0.406	-0.2	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	85	0.00
64 T	Tetrachloroethene	0.351	0.354	-0.9	95	0.00
65 PM	Chlorobenzene	1.079	1.099	-1.9	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.405	-5.2	96	0.00
67 C	Ethyl Benzene	1.924	2.009	-4.4#	92	0.00
68 T	m/p-Xylenes	0.733	0.773	-5.5	93	0.00
69 T	o-Xylene	0.656	0.704	-7.3	94	0.00
70 T	Styrene	1.134	1.217	-7.3	92	0.00
71 P	Bromoform	0.195	0.213	-9.2	99	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
73 T	Isopropylbenzene	3.711	3.898	-5.0	96	0.00
74 T	N-aryl acetate	0.842	0.889	-5.6	95	0.00
75 P	1,1,2,2-Tetrachloroethane	0.689	0.728	-5.7	100	0.00
76 T	1,2,3-Trichloropropane	0.421	0.542	-28.7#	115	0.00
77 T	Bromobenzene	0.849	0.895	-5.4	95	0.00
78 T	n-propylbenzene	4.613	4.882	-5.8	96	0.00
79 T	2-Chlorotoluene	2.606	2.680	-2.8	93	0.00
80 T	1,3,5-Trimethylbenzene	3.082	3.272	-6.2	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.205	0.222	-8.3	100	0.00
82 T	4-Chlorotoluene	2.739	2.864	-4.6	95	0.00
83 T	tert-Butylbenzene	2.535	2.733	-7.8	97	0.00
84 T	1,2,4-Trimethylbenzene	3.091	3.288	-6.4	95	0.00
85 T	sec-Butylbenzene	3.709	3.948	-6.4	97	0.00
86 T	p-Isopropyltoluene	3.264	3.546	-8.6	98	0.00
87 T	1,3-Dichlorobenzene	1.688	1.735	-2.8	96	0.00
88 T	1,4-Dichlorobenzene	1.667	1.692	-1.5	95	0.00
89 T	n-Butylbenzene	3.238	3.428	-5.9	96	0.00
90 T	Hexachloroethane	0.667	0.684	-2.5	96	0.00
91 T	1,2-Dichlorobenzene	1.421	1.503	-5.8	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.108	0.109	-0.9	105	0.00
93 T	1,2,4-Trichlorobenzene	0.844	0.910	-7.8	101	0.00
94 T	Hexachlorobutadiene	0.494	0.502	-1.6	99	0.00
95 T	Naphthalene	1.491	1.712	-14.8	103	0.00
96 T	1,2,3-Trichlorobenzene	0.729	0.787	-8.0	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122420\
Data File : VD067950.D
Acq On : 24 Dec 2020 17:05
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Dec 25 00:16:35 2020
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D122220S.M
Quant Title : Sw846 8260
QLast Update : Wed Dec 23 07:34:16 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 6